



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	David J. Zeis				
Location:	Cavalier County Sheriff's Office - Langdon				
Intoxilyzer® Serial Number:	80-004937				
Gas Lot Number:	659358	Gas Cylinder Number:	22	Expiration Date:	09/03/2011
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.079 %					
Test 2: 0.078 %					
Test 3: 0.078 %					
(ATTACH TEST RECORD)					

David J. Zeis
Chemical Test Operator Signature

04/01/11
Date

Deb Kashner
Reviewed By

14 Apr 11
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = LANG 8164.13.00 06/09
04/01/2011 13:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.079	13:30
03 Room Air	0.000	13:30
04 Std. Gas	0.078	13:31
05 Room Air	0.000	13:31
06 Std. Gas	0.078	13:31
07 Room Air	0.000	13:32

Lot No = 659358
Cyl No = 22
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZELS

Remarks:

Form 106-I8000

80-604937
p. 59 of 59
JFM