



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lori Staten</u>		
Location: <u>Mandan PD</u>		
Intoxilyzer® Serial Number: <u>80-004939</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>40</u>	Expiration Date: <u>09-03-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. <u>082</u> %		
Test 3: 0. <u>082</u> %		
(ATTACH TEST RECORD)		

[Signature]  
Chemical Test Operator Signature

04-01-11  
Date

Deb Kashur  
Reviewed By

14 Apr 11  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MAND      8164.13.00 06/09  
04/01/2011      13:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.081	13:57
03 Room Air	0.000	13:57
04 Std. Gas	0.082	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.082	13:58
07 Room Air	0.000	13:59

Lot No = 659358  
Cyl No = 40  
Exp Date = 09/03/2011  
County = 30      Oper No. = 068019

  
Operator Signature  
LORI E. FLATEN

Remarks:

Form 106-I8000

80-004939  
P. 52 of 52  
YH