

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004947

Location: TOXL

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>✓</u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5134</u> Lot No. <u>✓</u> Rep AC <u>0.294</u>  | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>✓</u> Rep AC <u>0.10 AC + acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. 1541008041 Cyl No. 12 Exp. Date 07/01/2012

Test 1 0.079 Test 4 0.079 Test 7 0.080

Test 2 0.079 Test 5 0.079 Test 8 0.079

Test 3 0.080 Test 6 0.080 Test 9 0.079

Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 459 psi Reg 450 psi

M. Remarks/Maintenance record: Annual inspection

01 Apr 2011  
Date

Deb Kashur  
Field Inspector's Signature

1 Apr 2011  
Date

Margaret A. Blum  
Reviewed by

80-004947  
P. 60 of 68  
RM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
04/01/2011      10:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:42
02 Std. Sol.	0.019	10:43
03 Room Air	0.000	10:43
04 Std. Sol.	0.019	10:44
05 Room Air	0.000	10:45
06 Std. Sol.	0.019	10:45
07 Room Air	0.000	10:46

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Low AC 0.021 AC

Form 106-I8000

80-004947  
P. 62 of 68  
AM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004947  
Location = TOXL                    8164.13.00 06/09  
04/01/2011                            10:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:52
02 Std. Sol.	0.285	10:53
03 Room Air	0.000	10:54
04 Std. Sol.	0.288	10:54
05 Room Air	0.000	10:55
06 Std. Sol.	0.290	10:56
07 Room Air	0.000	10:56

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Keshur

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.294 AC*

Form 106-18000

80-004947  
P. 63 of 68  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004947  
Location = TOXL                    8164.13.00 06/09  
04/01/2011                            11:10

Test	AC	Time
01 Room Air	0.000	11:11
02 *Subject Test	INT*	11:12
03 Room Air	0.000	11:13

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 12  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10 AC + acetone*

Form 106-I8000

80-004947  
P. 6/1/08  
AM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004947  
Location = TOXL                8164.13.00 06/09  
04/01/2011                        11:08

Test	AC	Time
01 Diagnostic	OK	11:09
02 Room Air	RFI*	11:09
03 Room Air	0.000	11:09

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 12  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004947  
P. 65 of 68  
JAN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004947  
Location = TOXL                    8164.13.00 06/09  
04/01/2011                            10:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:09
02 Std. Gas	0.079	10:10
03 Room Air	0.000	10:10
04 Std. Gas	0.079	10:11
05 Room Air	0.000	10:11
06 Std. Gas	0.080	10:12
07 Room Air	0.000	10:12

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08                            Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004947  
P. 66 of 68  
AK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
04/01/2011      10:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:13
02 Std. Gas	0.079	10:13
03 Room Air	0.000	10:14
04 Std. Gas	0.079	10:14
05 Room Air	0.000	10:15
06 Std. Gas	0.080	10:15
07 Room Air	0.000	10:16

Lot No = 15410080A1

Cyl No = 12

Exp Date = 07/01/2012

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004947  
P.67 of 68  
JK



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = TOXL      8164.13.00 06/09  
04/01/2011      10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.080	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.079	10:20
05 Room Air	0.000	10:20
06 Std. Gas	0.079	10:20
07 Room Air	0.000	10:21

Lot No = 15410080A1

Cyl No = 12

Exp Date = 07/01/2012

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080A*

Form 106-I8000

*80-004947  
P. 68 of 68  
JK*