

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003066

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>12210</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3803</u> Lot No. <u>12050</u> Rep AC <u>0.300</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>—</u> Rep AC <u>0.100 AC + acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>23411080 A1</u>	Cyl No. <u>15</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.079</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 780 psi Reg 760 psi

M. Remarks/Maintenance record: \*Annual inspection  
New external battery, recalibrated gas regulator

01 Apr 2013  
Date

Deb Kashner  
Field Inspector's Signature

01 APR 2013  
Date

Charles E. Edm  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003066  
Location = TOXL                    8164.13.00 06/09  
04/01/2013                            10:25

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            15  
Standard Expiration?        10/01/2013  
Oper No?                    777777

Flow Cal. Date:            07/29/2010  
Slope                        806  
Intercept                  -895033

IR Calibration Date:        07/29/2010

	3um	9um
0th Coef(*100):	-14075	-5808
1st Coef(*100):	272113	138204
2nd Coef(*100):	1637	986
H2O adj(mg/l*10k):	289	148

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
04/01/2013      10:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Sol.	0.022	10:29
03 Room Air	0.000	10:29
04 Std. Sol.	0.021	10:30
05 Room Air	0.000	10:31
06 Std. Sol.	0.021	10:31
07 Room Air	0.000	10:32

08 Sim Temp = 34.0°C

Simul Ser No = DR5132

Std Sol No = 12210

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003066  
Location = TOXL                    8164.13.00 06/09  
04/01/2013                            10:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:34
02 Std. Sol.	0.299	10:35
03 Room Air	0.000	10:35
04 Std. Sol.	0.299	10:36
05 Room Air	0.000	10:37
06 Std. Sol.	0.298	10:38
07 Room Air	0.000	10:38

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 12050

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003066  
Location = TOXL                8164.13.00 06/09  
04/01/2013                        10:40

Test	AC	Time
01 Room Air	0.000	10:40
02 *Subject Test	INT*	10:41
03 Room Air	0.000	10:41

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 15  
Expiration Date = 10/01/2013  
County = 08                        Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003066  
Location = TOXL                    8164.13.00 06/09  
04/01/2013                            10:43

Test	AC	Time
01 Diagnostic	OK	10:44
02 Room Air	RFI*	10:44
03 Room Air	0.000	10:44

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 15  
Expiration Date = 10/01/2013  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003066  
Location = TOXL                 8164.13.00 06/09  
04/01/2013                        10:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Gas	0.079	10:47
03 Room Air	0.000	10:48
04 Std. Gas	0.079	10:48
05 Room Air	0.000	10:49
06 Std. Gas	0.080	10:49
07 Room Air	0.000	10:50

Lot No = 23411080A1  
Cyl No = 15  
Exp Date = 10/01/2013  
County = 08                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
04/01/2013      10:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:50
02 Std. Gas	0.079	10:51
03 Room Air	0.000	10:51
04 Std. Gas	0.080	10:52
05 Room Air	0.000	10:52
06 Std. Gas	0.079	10:53
07 Room Air	0.000	10:53

Lot No = 23411080A1  
Cyl No = 15  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
Calibration check 0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = TOXL      8164.13.00 06/09  
04/01/2013      10:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:54
02 Std. Gas	0.080	10:54
03 Room Air	0.000	10:55
04 Std. Gas	0.079	10:55
05 Room Air	0.000	10:56
06 Std. Gas	0.080	10:56
07 Room Air	0.000	10:57

Lot No = 23411080A1

Cyl No = 15

Exp Date = 10/01/2013

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000