

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004934 Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5189 Lot No. 13300 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS#2 Rep AC 0.100AC+0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 034 Exp. Date 08/01/2014

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.081</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 254 psi Reg ²⁵⁰275 ^{CEC} psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

01 APR 14
Date

Charles E. Eden
Field Inspector's Signature

01 Apr 2014
Date

Deb Kasher
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 14:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 34
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 01/20/2010
Slope 698
Intercept -694859

IR Calibration Date: 04/03/2013

	3um	9um
0th Coef(*100):	-12034	-14742
1st Coef(*100):	249948	127320
2nd Coef(*100):	3434	1562
H2O adj(mg/l*10k):	561	610

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 14:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:48
02 Std. Sol.	0.023	14:49
03 Room Air	0.000	14:49
04 Std. Sol.	0.023	14:50
05 Room Air	0.000	14:50
06 Std. Sol.	0.023	14:51
07 Room Air	0.000	14:52

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 12111
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LOW AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 14:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:53
02 Std. Sol.	0.306	14:54
03 Room Air	0.000	14:54
04 Std. Sol.	0.308	14:55
05 Room Air	0.000	14:56
06 Std. Sol.	0.308	14:56
07 Room Air	0.000	14:57

08 Sim Temp = 34.0°C

Simul Ser No = DR5189
Std Sol No = 13300
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 15:00

Test	AC	Time
01 Room Air	0.000	15:01
02 *Subject Test	INT*	15:01
03 Room Air	0.000	15:02

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 34
Expiration Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK
0.100AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 15:02

Test	AC	Time
01 Diagnostic	OK	15:03
02 Room Air	RFI*	15:03
03 Room Air	0.000	15:04

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 34
Expiration Date = 08/01/2014
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 15:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:05
02 Std. Gas	0.080	15:05
03 Room Air	0.000	15:05
04 Std. Gas	0.080	15:06
05 Room Air	0.000	15:06
06 Std. Gas	0.081	15:07
07 Room Air	0.000	15:07

Lot No = 16512080A2
Cyl No = 34
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.080 AC
CALIBRATION CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 15:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:09
02 Std. Gas	0.080	15:09
03 Room Air	0.000	15:10
04 Std. Gas	0.080	15:10
05 Room Air	0.000	15:10
06 Std. Gas	0.080	15:11
07 Room Air	0.000	15:11

Lot No = 16512080A2
Cyl No = 34
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.080 AC
CALIBRATION CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
04/01/2014 15:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:12
02 Std. Gas	0.080	15:13
03 Room Air	0.000	15:13
04 Std. Gas	0.080	15:13
05 Room Air	0.000	15:14
06 Std. Gas	0.080	15:14
07 Room Air	0.000	15:15

Lot No = 16512080A2
Cyl No = 34
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

0.080AC

CALIBRATION CHECK

Form 106-I8000