

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004943

Location: Mountain Co SO - Stanley-31

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) Sim SN <u>DR5132</u> Lot No. <u> </u> Rep AC <u>0.021</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.) Sim SN <u>DR5134</u> Lot No. <u> </u> Rep AC <u>0.294</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) Sim SN <u>DR3846</u> Lot No. <u> </u> Rep AC <u>0.10 AC + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.) | |

Lot No. 709352 Cyl No. 34 Exp. Date 12/23/2011

Test 1 0.080 Test 4 0.080 Test 7 0.079

Test 2 0.080 Test 5 0.079 Test 8 0.079

Test 3 0.079 Test 6 0.080 Test 9 0.079

Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 705 psi Reg 700 psi

M. Remarks/Maintenance record: Annual inspection

04 Apr 2011
Date

Deb Kashner
Field Inspector's Signature

4-14-11
Date

[Signature]
Reviewed by

80-004943
p. 71 of 79
KPM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 34
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 693
Intercept -665387

IR Calibration Date: 01/18/2010
 3um 9um

0th Coef(*100): -20351 -24446
1st Coef(*100): 248542 131200
2nd Coef(*100): 2180 1190
H2O adj(mg/l*10k): 594 450

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:29

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 09:30 |
| 02 Std. Sol. | 0.017 | 09:31 |
| 03 Room Air | 0.000 | 09:31 |
| 04 Std. Sol. | 0.017 | 09:32 |
| 05 Room Air | 0.000 | 09:33 |
| 06 Std. Sol. | 0.017 | 09:33 |
| 07 Room Air | 0.000 | 09:34 |

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 31

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.021 AC

Form 106-I8000

80-004943
p. 73 of 79
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:35

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 09:36 |
| 02 Std. Sol. | 0.282 | 09:37 |
| 03 Room Air | 0.000 | 09:37 |
| 04 Std. Sol. | 0.285 | 09:38 |
| 05 Room Air | 0.000 | 09:38 |
| 06 Std. Sol. | 0.285 | 09:39 |
| 07 Room Air | 0.000 | 09:40 |

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 31

Oper No. = 777777

Deb Kasher

Operator Signature

DEB KASHUR

Remarks:

Linearity test 0.294AC

Form 106-I8000

80-004943
P. 74479
SM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:25

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 09:26 |
| 02 *Subject Test | INT* | 09:26 |
| 03 Room Air | 0.000 | 09:27 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 34
Expiration Date = 12/23/2011
County = 31 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Interferent detected 0.10Ac + acetone

Form 106-I8000

80-004943
P. 75 of 79
AK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:09

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 09:10 |
| 02 Room Air | RFI* | 09:10 |
| 03 Room Air | 0.000 | 09:11 |

*Invalid Test
Inhibited - RFI


Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 34
Expiration Date = 12/23/2011
County = 31 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:



Form 106-I8000

80-004943
P. 76 of 79
JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:11

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:12 |
| 02 Std. Gas | 0.080 | 09:12 |
| 03 Room Air | 0.000 | 09:13 |
| 04 Std. Gas | 0.080 | 09:13 |
| 05 Room Air | 0.000 | 09:13 |
| 06 Std. Gas | 0.079 | 09:14 |
| 07 Room Air | 0.000 | 09:14 |

Lot No = 709352
Cyl No = 34
Exp Date = 12/23/2011
County = 31 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004943
P. 770/79
JK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:16

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:17 |
| 02 Std. Gas | 0.080 | 09:17 |
| 03 Room Air | 0.000 | 09:18 |
| 04 Std. Gas | 0.079 | 09:18 |
| 05 Room Air | 0.000 | 09:19 |
| 06 Std. Gas | 0.080 | 09:19 |
| 07 Room Air | 0.000 | 09:19 |

Lot No = 709352
Cyl No = 34
Exp Date = 12/23/2011
County = 31 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004943
p. 78 of 79
VM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
04/04/2011 09:20

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:20 |
| 02 Std. Gas | 0.079 | 09:20 |
| 03 Room Air | 0.000 | 09:21 |
| 04 Std. Gas | 0.079 | 09:21 |
| 05 Room Air | 0.000 | 09:22 |
| 06 Std. Gas | 0.079 | 09:22 |
| 07 Room Air | 0.000 | 09:23 |

Lot No = 709352
Cyl No = 34
Exp Date = 12/23/2011
County = 31 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

*80-004943
p. 79 of 79
FM*