

INTOXILYZER 8000 INSPECTION

Serial No.: 80-005361 Location: TOXU

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
  - B. Is a three-pronged grounded outlet used? ✓
  - C. Is the breath tube heated? ✓
  - D. Is the diagnostic check complete? ✓
  - E. Is the time, date, and year correct (re-set if necessary)? ✓
  - F. Print test. (Attach test record.) ✓
  - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR5132 Lot No. 10290 Rep AC 0.020 ✓
  - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR3803 Lot No. 11110 Rep AC 0.300 ✓
  - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR3846 Lot No.      Rep AC 0.100 ✓
  - J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
  - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. <u>23411080A1</u>	Cyl No. <u>16</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>

Average                     

L. Tank pressure, Level 3 Func D Sub G P: 937 psi Reg 940 psi

M. Remarks/Maintenance record: Annual inspection

04 Apr 2012  
Date                     

Debs Kashur  
Field Inspector's Signature

09 APR 12  
Date                     

                      
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      14:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:16
02 Std. Sol.	0.020	14:16
03 Room Air	0.000	14:17
04 Std. Sol.	0.021	14:18
05 Room Air	0.000	14:18
06 Std. Sol.	0.021	14:19
07 Room Air	0.000	14:19

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Low Ac      0.020 Ac

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      14:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Sol.	0.304	14:49
03 Room Air	0.000	14:50
04 Std. Sol.	0.303	14:51
05 Room Air	0.000	14:51
06 Std. Sol.	0.302	14:52
07 Room Air	0.000	14:53


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

Linearity test      0.030Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      15:04

Test	AC	Time
01 Room Air	0.000	15:05
02 *Subject Test	INT*	15:05
03 Room Air	0.000	15:06

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 16  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      15:06

Test	AC	Time
01 Diagnostic	OK	15:07
02 Room Air	RFI*	15:07
03 Room Air	0.000	15:08

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 16  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      13:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:25
02 Std. Gas	0.081	13:26
03 Room Air	0.000	13:26
04 Std. Gas	0.080	13:26
05 Room Air	0.000	13:27
06 Std. Gas	0.080	13:27
07 Room Air	0.000	13:28

Lot No = 23411080A1  
Cyl No = 16  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kashur*  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      13:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:41
02 Std. Gas	0.080	13:42
03 Room Air	0.000	13:42
04 Std. Gas	0.081	13:43
05 Room Air	0.000	13:43
06 Std. Gas	0.080	13:44
07 Room Air	0.000	13:44

Lot No = 23411080A1  
Cyl No = 16  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005361  
Location = TOXL      8164.13.00 06/09  
04/04/2012      13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Gas	0.081	13:49
03 Room Air	0.000	13:49
04 Std. Gas	0.080	13:50
05 Room Air	0.000	13:50
06 Std. Gas	0.081	13:50
07 Room Air	0.000	13:51

Lot No = 23411080A1  
Cyl No = 16  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC