



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

|   |  |                                |
|---|--|--------------------------------|
| Chemical Test Operator Name: <u>Officer Pete Nielsen</u>      |  |                                |
| Location: <u>West Fargo Police Department</u>                 |  |                                |
| Intoxilyzer® Serial Number: <u>80-003059</u>                  |  |                                |
| Gas Lot Number: <u>659357</u>                                 | Gas Cylinder Number: <u>10</u>                           | Expiration Date: <u>9-3-11</u> |
| <input type="checkbox"/> New Cylinder Installed               | <input checked="" type="checkbox"/> Cylinder Not Changed |                                |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) |  |                                |
| Test 1: 0. <u>081</u> %                                       |  |                                |
| Test 2: 0. <u>081</u> %                                       |  |                                |
| Test 3: 0. <u>081</u> %                                       |  |                                |
| (ATTACH TEST RECORD)  |  |                                |

Chemical Test Operator Signature

4-5-10  
Date

Deb Kashur  
Reviewed By

07 Apr 2010  
Date

80-003059  
80/81 FN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = WFPD      8164.13.00 06/09  
04/05/2010      10:05

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:06 |
| 02 Std. Gas | 0.081 | 10:06 |
| 03 Room Air | 0.000 | 10:07 |
| 04 Std. Gas | 0.081 | 10:07 |
| 05 Room Air | 0.000 | 10:08 |
| 06 Std. Gas | 0.081 | 10:08 |
| 07 Room Air | 0.000 | 10:09 |

Lot No = 659357  
Cyl No = 10  
Exp Date = 09/03/2011  
County = 09      Oper No. = 039617



Operator Signature  
PETER D. NIELSEN

Remarks:

Form 106-I8000

80-003059  
8/1/81 *PN*