

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson JEC

Intoxilyzer Serial No.: 80-003066

Lot Number: 165 9358

Cylinder Number: 23

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082 %

Test 2: 0.082 %

Test 3: 0.082 %

(ATTACH TEST RECORD)

DM Wallace  
Chemical Test Operator Signature

4-5-10  
Date

Deb Kashur  
Reviewed By

07 Apr 2010  
Date

80-003066  
P.82/83 KM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = DKPD      8164.13.00 06/09  
04/05/2010      11:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:28
02 Std. Gas	0.082	11:28
03 Room Air	0.000	11:29
04 Std. Gas	0.082	11:29
05 Room Air	0.000	11:30
06 Std. Gas	0.082	11:30
07 Room Air	0.000	11:31

Lot No = 659358  
Cyl No = 23  
Exp Date = 09/03/2011  
County = 45      Oper No. = 068332

  
Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-003066  
P. 83/83 KM