

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

### ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Berni Vannatta

Location: Mohl

Intoxilyzer Serial No.: 86-004198

Lot Number: 659358

Cylinder Number: 27

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 079 %

Test 2: 0. 079 %

Test 3: 0. 078 %

(ATTACH TEST RECORD)

Berni Vannatta  
Chemical Test Operator Signature

4-5-2010  
Date

Debi Kashur  
Reviewed By

07 Apr 2010  
Date

80-004198  
8/83 PL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = MOHL      8164.13.00 06/09  
04/05/2010      13:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.079	13:52
03 Room Air	0.000	13:52
04 Std. Gas	0.079	13:53
05 Room Air	0.000	13:53
06 Std. Gas	0.078	13:53
07 Room Air	0.000	13:54

Lot No = 659358

Cyl No = 27

Exp Date = 09/03/2011

County = 38

Oper No. = 039909



Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004198  
8/2/83 YAH