

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004204

Location: ToxL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>—</u> Rep AC <u>0.021</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>—</u> Rep AC <u>0.294</u>                               | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3452</u> Lot No. <u>—</u> Rep AC <u>0.10</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>IS410080A1</u>	Cyl No. <u>14</u>	Exp. Date <u>7-1-12</u>
Test 1 <u>0.083</u>	Test 4 <u>0.083</u>	Test 7 <u>0.083</u>
Test 2 <u>0.083</u>	Test 5 <u>0.084</u>	Test 8 <u>0.082</u>
Test 3 <u>0.083</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.083</u>		

L. Tank pressure, Level 3 Func D Sub G  
P: 2 psi Reg 2 psi

M. Remarks/Maintenance record: Annual Inspection  
4-5-11  
Date  
05 Apr 2011  
Date

Deb Kashner  
Field Inspector's Signature  
Deb Kashner  
Reviewed by

80-004204  
p. 92 of 100  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/05/2011                      08:49

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            15410080A1  
Standard Cyl #?            14  
Standard Expiration?       07/01/2012  
Oper No?                    555555

Flow Cal. Date:            07/13/2010  
Slope                      750  
Intercept                  -815939

IR Calibration Date:        07/13/2010  
                                    3um                      9um

-----  
0th Coef(\*100):            -24262                    -22939  
1st Coef(\*100):            266889                   136772  
2nd Coef(\*100):            1535                      960  
H2O adj(mg/l\*10k):        459                        425

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ASHLEY AYD

Remarks: *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/05/2011                        08:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:52
02 Std. Sol.	0.019	08:52
03 Room Air	0.000	08:53
04 Std. Sol.	0.019	08:54
05 Room Air	0.000	08:54
06 Std. Sol.	0.019	08:55
07 Room Air	0.000	08:55

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature  
ASHLEY AYD

Remarks: Low AC 0.021 AC

Form 106-I8000

80-004204  
p. 94 of 100  
vrm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/05/2011                        08:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:57
02 Std. Sol.	0.299	08:58
03 Room Air	0.000	08:59
04 Std. Sol.	0.301	08:59
05 Room Air	0.000	09:00
06 Std. Sol.	0.302	09:01
07 Room Air	0.000	09:01


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature  
ASHLEY AYD

Remarks: High AC 0.294 AC

Form 106-I8000

80-004204  
P. 95 of 100  
FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004204  
Location = TOXL                              8164.13.00 06/09  
04/05/2011    09:03

Test	AC	Time
01 Room Air	0.000	09:04
02 *Subject Test	INT*	09:05
03 Room Air	0.000	09:05

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                              Weight = 150  
Test = DUI    Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 0  
Std Sol No = 0  
County = 08                                      Oper No. = 555555



Operator Signature  
ASHLEY AYD

Remarks: Interferent Acetone <sup>0.05%</sup> + 0.10 Alcohol

Form 106-I8000

80-004204  
P. 9/2/100  
10/22

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
04/05/2011                            09:07

Test	AC	Time
01 Diagnostic	OK	09:08
02 Room Air	RFI*	09:09
03 Room Air	0.000	09:09

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 0  
Std Sol No = 0  
County = 08                            Oper No. = 555555

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ASHLEY AYD

Remarks: RFI

Form 106-I8000

80-004204  
p. 97 of 100  
VFM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/05/2011                        09:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:13
02 Std. Gas	0.083	09:13
03 Room Air	0.000	09:14
04 Std. Gas	0.083	09:14
05 Room Air	0.000	09:15
06 Std. Gas	0.083	09:15
07 Room Air	0.000	09:15

Lot No = 15410080A1  
Cyl No = 14  
Exp Date = 07/01/2012  
County = 08                            Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: *Calibration check 0.080 AC*

Form 106-I8000

*80-004204  
P. 98 of 100  
VAM*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
04/05/2011                            09:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:18
02 Std. Gas	0.083	09:18
03 Room Air	0.000	09:19
04 Std. Gas	0.084	09:19
05 Room Air	0.000	09:19
06 Std. Gas	0.082	09:20
07 Room Air	0.000	09:20

Lot No = 15410080A1  
Cyl No = 14  
Exp Date = 07/01/2012  
County = 08                            Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-004204  
P. 99 of 100  
FM




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
04/05/2011                            09:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Gas	0.083	09:22
03 Room Air	0.000	09:23
04 Std. Gas	0.082	09:23
05 Room Air	0.000	09:24
06 Std. Gas	0.082	09:24
07 Room Air	0.000	09:25

Lot No = 15410080A1  
Cyl No = 14  
Exp Date = 07/01/2012  
County = 08                            Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-004204  
P.100 of 100  
JFM