

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003058

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. — Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. — Rep AC 0.294 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3452 Lot No. — Rep AC 0.10 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 15410080A1 Cyl No. 12 Exp. Date 7-1-12
Test 1 0.079 Test 4 0.080 Test 7 0.080
Test 2 0.079 Test 5 0.079 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 308 psi Reg 300 psi

M. Remarks/Maintenance record: Annual Inspection

4-6-11
Date

Deb Kashur
Field Inspector's Signature

06 Apr 2011
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 12
Standard Expiration? 07/01/2012
Oper No? 555555

Flow Cal. Date: 04/07/2008
Slope 698
Intercept -678970

IR Calibration Date: 06/05/2008
 3um 9um

0th Coef(*100):	-18867	-13178
1st Coef(*100):	274682	135660
2nd Coef(*100):	1139	678
H2O adj(mg/l*10k):	434	216

***** Printer Test End *****



Operator Signature
ASHLEY AYD

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Sol.	0.019	13:36
03 Room Air	0.000	13:36
04 Std. Sol.	0.019	13:37
05 Room Air	0.000	13:37
06 Std. Sol.	0.020	13:38
07 Room Air	0.000	13:39

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Low AC 0.021 AC

Form 106-I8000

80-003058
P.954/101
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Sol.	0.294	13:44
03 Room Air	0.000	13:45
04 Std. Sol.	0.296	13:45
05 Room Air	0.000	13:46
06 Std. Sol.	0.297	13:47
07 Room Air	0.000	13:47


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: High AC 0.294 AC

Form 106-I8000

80-003058
P. 96 of 101
EM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:48

Test	AC	Time
01 Room Air	0.000	13:49
02 *Subject Test	INT*	13:49
03 Room Air	0.000	13:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 12
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Interferent check 0.10 AC + Acetone*

Form 106-I8000

80-003058
P. 97 of 101
RM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:51

Test	AC	Time
01 Diagnostic	OK	13:52
02 Room Air	RFI*	13:52
03 Room Air	0.000	13:52

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 12
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ASHLEY AYD

Remarks: RFI

Form 106-I8000

80-003058
P. 98 of 101
sam


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:53
02 Std. Gas	0.079	13:54
03 Room Air	0.000	13:54
04 Std. Gas	0.079	13:55
05 Room Air	0.000	13:55
06 Std. Gas	0.079	13:55
07 Room Air	0.000	13:56

Lot No = 15410080A1
Cyl No = 12
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-003058
P. 99 of 101
JAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:57
02 Std. Gas	0.080	13:57
03 Room Air	0.000	13:58
04 Std. Gas	0.079	13:58
05 Room Air	0.000	13:58
06 Std. Gas	0.079	13:59
07 Room Air	0.000	13:59

Lot No = 15410080A1
Cyl No = 12
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-003058
P. 100 of 101
vsm


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
04/06/2011 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:00
02 Std. Gas	0.080	14:00
03 Room Air	0.000	14:01
04 Std. Gas	0.079	14:01
05 Room Air	0.000	14:02
06 Std. Gas	0.079	14:02
07 Room Air	0.000	14:03

Lot No = 15410080A1
Cyl No = 12
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: ~~Calibration~~ Calibration check
0.080 AC

Form 106-I8000

80-003058
P. 101 of 101
JAM