



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Matthew Anderson</u>		
Location: <u>Richland County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004206</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>32</u>	Expiration Date: <u>7/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.080</u> % Test 2: <u>0.080</u> % Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Matthew Anderson
 Chemical Test Operator Signature

4/10/2011
 Date

Deb Kashner
 Reviewed By

14 Apr 11
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
04/06/2011 20:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:44
02 Std. Gas	0.080	20:44
03 Room Air	0.000	20:45
04 Std. Gas	0.080	20:45
05 Room Air	0.000	20:45
06 Std. Gas	0.079	20:46
07 Room Air	0.000	20:46


Lot No = 15410080A1

Cyl No = 32

Exp Date = 07/01/2012

County = 39

Oper No. = 020401



Operator Signature
MATTHEW W. ANDERSON

Remarks:

Form 106-I8000

80-004206
p. 117 of 117
FM