

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005948

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 939 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: IC66 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0262108A1 Cyl #: 4 Exp. Date: 3/5/2024

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: Annual Inspection

Re placed o-ring on breath hose. Replaced o-ring on simulator return.

Instrument is acceptable to be used in the field.

Janelle PottsChiller
Breath Analyst Signature

4/6/2020
Date

[Signature]
Reviewed by

04/22/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = TOXL 8164.14.00 09/16
04/06/2022 10:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Sol.	0.020	10:32
03 Room Air	0.000	10:32
04 Std. Sol.	0.020	10:33
05 Room Air	0.000	10:33
06 Std. Sol.	0.020	10:34
07 Room Air	0.000	10:35

08 Sim Temp = 34.0°C

Simul Ser No = MP3060
Std Sol No = 202201A
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = TOXL 8164.14.00 09/16
04/06/2022 10:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:37
02 Std. Sol.	0.297	10:37
03 Room Air	0.000	10:38
04 Std. Sol.	0.303	10:39
05 Room Air	0.000	10:39
06 Std. Sol.	0.305	10:40
07 Room Air	0.000	10:40

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 202012A
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *High Ac 0.300*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = TOXL 8164.14.00 09/16
04/06/2022 10:48

Test	AC	Time
01 Room Air	0.000	10:50
02 *Subject Test	INT*	10:52
03 Room Air	0.000	10:53

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = TOXL 8164.14.00 09/16
04/06/2022 10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.080	10:57
03 Room Air	0.000	10:57
04 Std. Gas	0.079	10:58
05 Room Air	0.000	10:58
06 Std. Gas	0.080	10:59
07 Room Air	0.000	10:59

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1 -3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = TOXL 8164.14.00 09/16
04/06/2022 11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Gas	0.080	11:01
03 Room Air	0.000	11:01
04 Std. Gas	0.080	11:02
05 Room Air	0.000	11:02
06 Std. Gas	0.080	11:02
07 Room Air	0.000	11:03

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005948
Location = TOXL 8164.14.00 09/16
04/06/2022 11:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:04
02 Std. Gas	0.080	11:04
03 Room Air	0.000	11:05
04 Std. Gas	0.080	11:05
05 Room Air	0.000	11:06
06 Std. Gas	0.080	11:06
07 Room Air	0.000	11:07

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*