



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger # 258</u>		
Location: <u>UND - Police Dept.</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger # 258  
Chemical Test Operator Signature

04-07-2011  
Date

Deb Kashner  
Reviewed By

13 Apr 2011  
Date

80-004934  
P.69 of 70  
JFM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004934  
Location = UNDP                    8164.13.00 06/09  
04/07/2011                            09:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Gas	0.080	09:25
03 Room Air	0.000	09:25
04 Std. Gas	0.081	09:25
05 Room Air	0.000	09:26
06 Std. Gas	0.081	09:26
07 Room Air	0.000	09:27

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18                            Oper No. = 119409

*Tracy L. Meidinger #258*  
Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA 04-07-2011

80-004934  
P. 70 of 70  
SM