

**INTOXILYZER 8000 INSPECTION**

Serial No.: 80-004940

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR5144 Lot No. — Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR3847 Lot No. — Rep AC 0.294 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR3452 Lot No. — Rep AC 0.10 AC + Acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 15410080A1 Cyl No. 12 Exp. Date 7-1-12  
Test 1 0.079 Test 4 0.077 Test 7 0.079  
Test 2 0.079 Test 5 0.078 Test 8 0.079  
Test 3 0.079 Test 6 0.078 Test 9 0.078  
Average 0.078

L. Tank pressure, Level 3 Func D Sub G P: 253 psi Reg 253 psi

M. Remarks/Maintenance record: Gas Cylinder Regulator Recalibrated - Annual Inspection

4-7-11  
Date

[Signature]  
Field Inspector's Signature

07 Apr 2011  
Date

Deb Kashner  
Reviewed by

80-004940  
p. 43 of 51  
FM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004940  
Location = TOXL      8164.13.00 06/09  
04/07/2011      07:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:47
02 Std. Sol.	0.021	07:48
03 Room Air	0.000	07:48
04 Std. Sol.	0.021	07:49
05 Room Air	0.000	07:49
06 Std. Sol.	0.021	07:50
07 Room Air	0.000	07:51


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: Low AC 0.021 AC

Form 106-I8000

80-004940  
p. 45 of 51  
JAM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004940  
Location = TOXL      8164.13.00 06/09  
04/07/2011      08:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:13
02 Std. Sol.	0.300	08:14
03 Room Air	0.000	08:14
04 Std. Sol.	0.300	08:15
05 Room Air	0.000	08:15
06 Std. Sol.	0.300	08:16
07 Room Air	0.000	08:17


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: *Linearity test 0.294 AC*

Form 106-I8000

80-004940  
p. 46 of 51  
KTM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004940  
Location = TOXL                    8164.13.00 06/09  
04/07/2011                            08:19

Test	AC	Time
01 Room Air	0.000	08:19
02 *Subject Test	INT*	08:20
03 Room Air	0.000	08:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 0  
Std Sol No = 0  
County = 08                            Oper No. = 555555



Operator Signature  
ASHLEY AYD

Remarks: *Interferent check 0.10AC + Acetone*

Form 106-I8000

80-004940  
P. 47 of 51  
KTM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004940  
Location = TOXL                    8164.13.00 06/09  
04/07/2011                            08:25

Test	AC	Time
01 Diagnostic	OK	08:26
02 Room Air	RFI*	08:26
03 Room Air	0.000	08:26

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 0  
Std Sol No = 0  
County = 08                            Oper No. = 555555

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ASHLEY AYD

Remarks: RFI

Form 106-I8000

80-004940  
p. 48 of 51  
FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004940  
Location = TOXL      8164.13.00 06/09  
04/07/2011      08:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:28
02 Std. Gas	0.079	08:28
03 Room Air	0.000	08:29
04 Std. Gas	0.079	08:29
05 Room Air	0.000	08:30
06 Std. Gas	0.079	08:30
07 Room Air	0.000	08:31

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08      Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-004940  
P.49451  
KM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004940  
Location = TOXL      8164.13.00 06/09  
04/07/2011      08:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:32
02 Std. Gas	0.077	08:32
03 Room Air	0.000	08:33
04 Std. Gas	0.078	08:33
05 Room Air	0.000	08:34
06 Std. Gas	0.078	08:34
07 Room Air	0.000	08:34

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08      Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: Calibration check 0.000 AC

Form 106-I8000

80-004940  
p. 50 of 51  
RM




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004940  
Location = TOXL                    8164.13.00 06/09  
04/07/2011                            08:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:35
02 Std. Gas	0.079	08:36
03 Room Air	0.000	08:36
04 Std. Gas	0.079	08:36
05 Room Air	0.000	08:37
06 Std. Gas	0.078	08:37
07 Room Air	0.000	08:38

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08                            Oper No. = 555555

  
\_\_\_\_\_  
Operator Signature  
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-004940  
p. 51 of 51  
KM