

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004948

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. — Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. — Rep AC 0.294 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3452 Lot No. — Rep AC 0.10 AC + Acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 15410080A1 Cyl No. 39 Exp. Date 7-1-12
Test 1 0.080 Test 4 0.079 Test 7 0.080
Test 2 0.080 Test 5 0.080 Test 8 0.080
Test 3 0.079 Test 6 0.080 Test 9 0.080
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 252 psi Reg 250 psi

M. Remarks/Maintenance record: Annual Inspection

4-7-11
Date

[Signature]
Field Inspector's Signature

07 Apr 2011
Date

[Signature]
Reviewed by

80-004948
p. 39 of 47
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 08:53

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 39
Standard Expiration? 07/01/2012
Oper No? 555555

Flow Cal. Date: 07/14/2010
Slope 778
Intercept -795769

IR Calibration Date: 07/14/2010
 3um 9um

0th Coef(*100):	-31061	-21991
1st Coef(*100):	265097	133596
2nd Coef(*100):	802	1052
H2O adj (mg/l*10k):	675	417

***** Printer Test End *****



Operator Signature
ASHLEY AYD

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 08:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:55
02 Std. Sol.	0.019	08:56
03 Room Air	0.000	08:57
04 Std. Sol.	0.020	08:57
05 Room Air	0.000	08:58
06 Std. Sol.	0.020	08:58
07 Room Air	0.000	08:59


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Low AC 0.021 AC

Form 106-I8000

80-004948
p. 41 of 47
FA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 09:00


WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:01
02 Std. Sol.	0.298	09:02
03 Room Air	0.000	09:02
04 Std. Sol.	0.299	09:03
05 Room Air	0.000	09:03
06 Std. Sol.	0.299	09:04
07 Room Air	0.000	09:05

08 Sim Temp = 34.0°C

Simul Ser No = 0
Std Sol No = 0
County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Linearity test 0.294 AC*

Form 106-I8000

*80-004948
P. 42 of 47
JAM*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 09:07

Test	AC	Time
01 Room Air	0.000	09:07
02 *Subject Test	INT*	09:08
03 Room Air	0.000	09:09

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 39
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Interferent check 0.10 AC + Acetone*

Form 106-I8000

80-004948
P43 of 47
JCH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 09:09

Test	AC	Time
01 Diagnostic	OK	09:10
02 Room Air	RFI*	09:10
03 Room Air	0.000	09:11

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 39
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ASHLEY AYD

Remarks: RFI

Form 106-I8000

80-004948
p 44 of 47
JAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 09:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:12
02 Std. Gas	0.080	09:12
03 Room Air	0.000	09:13
04 Std. Gas	0.080	09:13
05 Room Air	0.000	09:14
06 Std. Gas	0.079	09:14
07 Room Air	0.000	09:14

Lot No = 15410080A1
Cyl No = 39
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-004948
p.45 of 47
KHL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 09:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:15
02 Std. Gas	0.079	09:16
03 Room Air	0.000	09:16
04 Std. Gas	0.080	09:17
05 Room Air	0.000	09:17
06 Std. Gas	0.080	09:17
07 Room Air	0.000	09:18

Lot No = 15410080A1
Cyl No = 39
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration Checks 0.080 AC

Form 106-I8000

80-004948
p. 46 of 47
JH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = TOXL 8164.13.00 06/09
04/07/2011 09:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:20
02 Std. Gas	0.080	09:20
03 Room Air	0.000	09:20
04 Std. Gas	0.080	09:21
05 Room Air	0.000	09:21
06 Std. Gas	0.080	09:22
07 Room Air	0.000	09:22

Lot No = 15410080A1
Cyl No = 39
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration check 0.080 AC

Form 106-I8000

80-004948
p 47 of 47
EM