



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Lt. Tracy L. Meidinger #258

Location: UND-PD

Intoxilyzer Serial No.: 80-004206

Lot Number: 659358

Cylinder Number: 31

Expiration Date: 09-03-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.084 %

Test 2: 0.083 %

Test 3: 0.084 %

(ATTACH TEST RECORD)

Lt. Tracy L. Meidinger #258
 Chemical Test Operator Signature

04-08-2010
 Date Cylinder Changed

Deb Kashner
 Reviewed By

13 Apr 2010
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004206
 7/3/11 KAL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = UNDP 8164.13.00 06/09
04/08/2010 12:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Gas	0.084	12:40
03 Room Air	0.000	12:41
04 Std. Gas	0.083	12:41
05 Room Air	0.000	12:42
06 Std. Gas	0.084	12:42
07 Room Air	0.000	12:43

Lot No = 659358
Cyl No = 31
Exp Date = 09/03/2011
County = 18 Oper No. = 119409


Operator Signature
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

AEA
4-08-10

80-004206
74/74 KM