



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Rude, Josh</u>		
Location: <u>Stuttsman Co.</u>		
Intoxilyzer® Serial Number: <u>80-004198</u>		
Gas Lot Number: <u>05911080A1</u>	Gas Cylinder Number: <u>21</u>	Expiration Date: <u>02-01-13</u>
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.077</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Josh Rude
Chemical Test Operator Signature

04-08-11
Date

Deb Kashner
Reviewed By

19 Apr 11
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = ~~TOXE~~ JAME 8164.13.00 06/09
04/08/2011 DK 19 April 18:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:40
02 Std. Gas	0.079	18:41
03 Room Air	0.000	18:41
04 Std. Gas	0.077	18:41
05 Room Air	0.000	18:42
06 Std. Gas	0.078	18:42
07 Room Air	0.000	18:43

Lot No = 05911080A1
Cyl No = 21
Exp Date = 02/01/2013
County = 47 Oper No. = 100414



Operator Signature
JOSH RUDE

Remarks:

Form 106-I8000

80-004198
P. 132 of 132
FM