



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Aaron R. Hummel

Location: GFSO

Intoxilyzer Serial No.: 80-003057

Lot Number: 659357

Cylinder Number: 26

Expiration Date: 9/3/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082 %

Test 2: 0.082 %

Test 3: 0.082 %

(ATTACH TEST RECORD)

[Signature]
 Chemical Test Operator Signature

4/9/10 - ACA Only
 Date Cylinder Changed

Debi Kashner
 Reviewed By

14 Apr 2010
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003057
 88/r9 AR


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = GFSO 8164.13.00 06/09
04/09/2010 14:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:53
02 Std. Gas	0.082	14:54
03 Room Air	0.000	14:54
04 Std. Gas	0.082	14:55
05 Room Air	0.000	14:55
06 Std. Gas	0.082	14:55
07 Room Air	0.000	14:56

Lot No = 659357
Cyl No = 26
Exp Date = 09/03/2011
County = 18 Oper No. = 049907



Operator Signature
AARON HUMMEL

Remarks:

ACA

Form 106-I8000

80-003057
8/8/09 FM