

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004186 Location: TOXL

- | | Check When Done |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) Sim SN <u>DR5132</u> Lot No. <u>10290</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.) Sim SN <u>DR3803</u> Lot No. <u>1116</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) Sim SN <u>DR3846</u> Lot No. <u>---</u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.) | |

| | | |
|---------------------------|---------------------|-----------------------------|
| Lot No. <u>23411080A1</u> | Cyl No. <u>16</u> | Exp. Date <u>10/01/2013</u> |
| Test 1 <u>0.081</u> | Test 4 <u>0.081</u> | Test 7 <u>0.081</u> |
| Test 2 <u>0.081</u> | Test 5 <u>0.081</u> | Test 8 <u>0.081</u> |
| Test 3 <u>0.080</u> | Test 6 <u>0.081</u> | Test 9 <u>0.081</u> |
| Average <u>0.081</u> | | |

L. Tank pressure, Level 3 Func D Sub G P: 850 psi Reg 850 psi

M. Remarks/Maintenance record: Annual inspection

09 Apr 12
Date

Deb Kashur
Field Inspector's Signature

09 APR 12
Date

Charles E. [Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 09:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 16
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 06/03/2009
Slope 703
Intercept -668211

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -12098 -31908
1st Coef(*100): 257966 136312
2nd Coef(*100): 2942 1009
H2O adj(mg/l*10k): 434 609

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 11:06

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:07 |
| 02 Std. Sol. | 0.018 | 11:07 |
| 03 Room Air | 0.000 | 11:08 |
| 04 Std. Sol. | 0.019 | 11:09 |
| 05 Room Air | 0.000 | 11:09 |
| 06 Std. Sol. | 0.020 | 11:10 |
| 07 Room Air | 0.000 | 11:10 |

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 11:01

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:01 |
| 02 Std. Sol. | 0.296 | 11:02 |
| 03 Room Air | 0.000 | 11:03 |
| 04 Std. Sol. | 0.297 | 11:04 |
| 05 Room Air | 0.000 | 11:04 |
| 06 Std. Sol. | 0.296 | 11:05 |
| 07 Room Air | 0.000 | 11:05 |

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 10:31

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 10:32 |
| 02 *Subject Test | INT* | 10:32 |
| 03 Room Air | 0.000 | 10:33 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 16
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 10:25

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 10:26 |
| 02 Room Air | RFI* | 10:26 |
| 03 Room Air | 0.000 | 10:27 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 16
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 09:33

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:33 |
| 02 Std. Gas | 0.081 | 09:34 |
| 03 Room Air | 0.000 | 09:34 |
| 04 Std. Gas | 0.081 | 09:35 |
| 05 Room Air | 0.000 | 09:35 |
| 06 Std. Gas | 0.080 | 09:35 |
| 07 Room Air | 0.000 | 09:36 |

Lot No = 23411080A1
Cyl No = 16
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 09:36

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:37 |
| 02 Std. Gas | 0.081 | 09:37 |
| 03 Room Air | 0.000 | 09:38 |
| 04 Std. Gas | 0.081 | 09:38 |
| 05 Room Air | 0.000 | 09:39 |
| 06 Std. Gas | 0.081 | 09:39 |
| 07 Room Air | 0.000 | 09:40 |

Lot No = 23411080A1
Cyl No = 16
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = TOXL 8164.13.00 06/09
04/09/2012 09:40

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:40 |
| 02 Std. Gas | 0.081 | 09:41 |
| 03 Room Air | 0.000 | 09:41 |
| 04 Std. Gas | 0.081 | 09:42 |
| 05 Room Air | 0.000 | 09:42 |
| 06 Std. Gas | 0.081 | 09:42 |
| 07 Room Air | 0.000 | 09:43 |

Lot No = 23411080A1
Cyl No = 16
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC