

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004191

Location: ToxL

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 12210 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 12050 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3377 Lot No. — Rep AC 0.10 AC + ACE ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>11611080A2</u>	Cyl No. <u>108</u>	Exp. Date <u>06/01/2013</u>
Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 270 psi Reg 275 psi

M. Remarks/Maintenance record: REPLACED EXTERNAL BATTERY; REPLACED O-RING SIM RETURN;
ANNUAL INSPECTION

08 APR 13
Date

Charles E. Edin
Field Inspector's Signature

09 Apr 13
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 15:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 108
Standard Expiration? 06/01/2013
Oper No? 666666

Flow Cal. Date: 01/28/2010
Slope 738
Intercept -777574

IR Calibration Date: 01/28/2010
3um 9um

0th Coef(*100):	-23128	-33609
1st Coef(*100):	272485	131724
2nd Coef(*100):	1490	1056
H2O adj(mg/l*10k):	432	614

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 15:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:54
02 Std. Sol.	0.020	15:54
03 Room Air	0.000	15:55
04 Std. Sol.	0.020	15:56
05 Room Air	0.000	15:56
06 Std. Sol.	0.020	15:57
07 Room Air	0.000	15:57

08 Sim Temp = 34.0°C

Simul Ser No = DR5132

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 16:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:01
02 Std. Sol.	0.297	16:02
03 Room Air	0.000	16:02
04 Std. Sol.	0.298	16:03
05 Room Air	0.000	16:04
06 Std. Sol.	0.296	16:04
07 Room Air	0.000	16:05


08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 12050

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 16:52

Test	AC	Time
01 Room Air	0.000	16:53
02 *Subject Test	INT*	16:53
03 Room Air	0.000	16:54

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVC098765ZXDE
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 108
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10 AC + ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 16:28

Test	AC	Time
01 Diagnostic	OK	16:29
02 Room Air	RFI*	16:29
03 Room Air	0.000	16:30

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOP12345
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 108
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 16:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:10
02 Std. Gas	0.081	16:11
03 Room Air	0.000	16:11
04 Std. Gas	0.081	16:12
05 Room Air	0.000	16:12
06 Std. Gas	0.081	16:13
07 Room Air	0.000	16:13

Lot No = 11611080A2
Cyl No = 108
Exp Date = 06/01/2013
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 EBS

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 16:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:16
02 Std. Gas	0.081	16:16
03 Room Air	0.000	16:17
04 Std. Gas	0.081	16:17
05 Room Air	0.000	16:18
06 Std. Gas	0.081	16:18
07 Room Air	0.000	16:19

Lot No = 11611080A2
Cyl No = 108
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 EBS


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/08/2013 16:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:21
02 Std. Gas	0.081	16:22
03 Room Air	0.000	16:22
04 Std. Gas	0.081	16:23
05 Room Air	0.000	16:23
06 Std. Gas	0.081	16:24
07 Room Air	0.000	16:24

Lot No = 11611080A2
Cyl No = 108
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080ERS