

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004939

Location: Morton Co Corr - Mandan

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 10290 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 11110 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3377 Lot No. Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 116 11080 A2 Cyl No. 82 Exp. Date 06/01/2013
Test 1 0.081 Test 4 0.081 Test 7 0.081
Test 2 0.081 Test 5 0.081 Test 8 0.081
Test 3 0.080 Test 6 0.081 Test 9 0.081
Average 0.081

L. Tank pressure, Level 3 Func D Sub G P: 488 psi Reg 500 psi

M. Remarks/Maintenance record: Annual inspection

10 Apr 12
Date

Deb Kashur
Field Inspector's Signature

10 APR 12
Date

Charles E. Ed
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:25

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 82
Standard Expiration? 06/01/2013
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 724
Intercept -779308

IR Calibration Date: 03/10/2010
 3um 9um

0th Coef(*100): -45908 -35439
1st Coef(*100): 263128 135479
2nd Coef(*100): 1658 1038
H2O adj(mg/l*10k): 832 527

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:41
02 Std. Sol.	0.018	13:42
03 Room Air	0.000	13:42
04 Std. Sol.	0.018	13:43
05 Room Air	0.000	13:44
06 Std. Sol.	0.018	13:44
07 Room Air	0.000	13:45


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 30

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:46
02 Std. Sol.	0.301	13:47
03 Room Air	0.000	13:48
04 Std. Sol.	0.298	13:49
05 Room Air	0.000	13:49
06 Std. Sol.	0.298	13:50
07 Room Air	0.000	13:51

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 30

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:51

Test	AC	Time
01 Room Air	0.000	13:53
02 *Subject Test	INT*	13:53
03 Room Air	0.000	13:54

*Invalid Test
Interferent Detected

Sub Name = TEST, TEST NA
Sub DOB = 08/27/1970
Sub Sex = Female Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Simul Ser No = 0
Std Sol No = 0
County = 30 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Interferent test 0.100 AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:54

Test	AC	Time
01 Diagnostic	OK	13:55
02 Room Air	RFI*	13:55
03 Room Air	0.000	13:56

*Invalid Test
Inhibited - RFI

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Gas	0.081	13:27
03 Room Air	0.000	13:28
04 Std. Gas	0.081	13:28
05 Room Air	0.000	13:29
06 Std. Gas	0.080	13:29
07 Room Air	0.000	13:30

Lot No = 11611080A2
Cyl No = 82
Exp Date = 06/01/2013
County = 30 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:31
02 Std. Gas	0.081	13:31
03 Room Air	0.000	13:32
04 Std. Gas	0.081	13:32
05 Room Air	0.000	13:33
06 Std. Gas	0.081	13:33
07 Room Air	0.000	13:33

Lot No = 11611080A2
Cyl No = 82
Exp Date = 06/01/2013
County = 30 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
04/10/2012 13:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:34
02 Std. Gas	0.081	13:35
03 Room Air	0.000	13:35
04 Std. Gas	0.081	13:35
05 Room Air	0.000	13:36
06 Std. Gas	0.081	13:36
07 Room Air	0.000	13:37

Lot No = 11611080A2
Cyl No = 82
Exp Date = 06/01/2013
County = 30 Oper No. = 777777

Deb Keshur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC