

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003071

Location: TOXL

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u>—</u> Rep AC <u>0.10 AC</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>43</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 895 psi Reg 900 psi

M. Remarks/Maintenance record: ^{CBE} REPLACED TANK SET REGULATOR; TANK SENSOR CAL;
REPLACED O-RING SIM RETURN; ANNUAL INSPECTION
10 APR 13
Date 10 Apr 13
Date

Charles E. Eden
Field Inspector's Signature
Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 10:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:10
02 Std. Sol.	0.021	10:10
03 Room Air	0.000	10:11
04 Std. Sol.	0.022	10:12
05 Room Air	0.000	10:12
06 Std. Sol.	0.022	10:13
07 Room Air	0.000	10:13

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 12210
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.020 AC
 Low AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 10:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:16
02 Std. Sol.	0.300	10:18
03 Room Air	0.000	10:18
04 Std. Sol.	0.301	10:19
05 Room Air	0.000	10:20
06 Std. Sol.	0.302	10:20
07 Room Air	0.000	10:21

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 10:04

Test	AC	Time
01 Room Air	0.000	10:05
02 *Subject Test	INT*	10:05
03 Room Air	0.000	10:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJHG09
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 43
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 10:01

Test	AC	Time
01 Diagnostic	OK	10:02
02 Room Air	RFI*	10:02
03 Room Air	0.000	10:03

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 43
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

~~RFI~~
CEE
Charles Eder

Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 11:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:14
02 Std. Gas	0.080	11:15
03 Room Air	0.000	11:15
04 Std. Gas	0.081	11:16
05 Room Air	0.000	11:16
06 Std. Gas	0.080	11:17
07 Room Air	0.000	11:17

Lot No = 23411080A1
Cyl No = 43
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION ^{CEE} ~~CHART~~ CHECK
0.080 E.B.S.


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 11:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.081	11:18
03 Room Air	0.000	11:19
04 Std. Gas	0.080	11:19
05 Room Air	0.000	11:20
06 Std. Gas	0.081	11:20
07 Room Air	0.000	11:21

Lot No = 23411080A1
Cyl No = 43
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 E.B.S.


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/10/2013 11:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.081	11:22
03 Room Air	0.000	11:22
04 Std. Gas	0.080	11:23
05 Room Air	0.000	11:23
06 Std. Gas	0.081	11:24
07 Room Air	0.000	11:24

Lot No = 23411080A1
Cyl No = 43
Exp Date = 10/01/2013
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 E.B.S.