

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004952 Location: TOXL

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 10290 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 11110 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. — Rep AC 0.100 ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080A1 Cyl No. 49 Exp. Date 10/01/2013
Test 1 0.080 Test 4 0.080 Test 7 0.081
Test 2 0.080 Test 5 0.080 Test 8 0.080
Test 3 0.080 Test 6 0.081 Test 9 0.080
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 885 psi Reg 890 psi

M. Remarks/Maintenance record: Annual inspection

11 Apr 12
Date

Deb Kashur
Field Inspector's Signature

12 APR 12
Date

Charles E. [Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/11/2012 11:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:10
02 Std. Sol.	0.017	11:11
03 Room Air	0.000	11:11
04 Std. Sol.	0.018	11:12
05 Room Air	0.000	11:12
06 Std. Sol.	0.017	11:13
07 Room Air	0.000	11:14

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/11/2012 11:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:23
02 Std. Sol.	0.293	11:26
03 Room Air	0.000	11:26
04 Std. Sol.	0.293	11:27
05 Room Air	0.000	11:27
06 Std. Sol.	0.292	11:28
07 Room Air	0.000	11:29

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/11/2012 11:42

Test	AC	Time
01 Room Air	0.000	11:43
02 *Subject Test	INT*	11:43
03 Room Air	0.000	11:44

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 0
Std Sol No = 0
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/11/2012 10:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:02
02 Std. Gas	0.080	10:02
03 Room Air	0.000	10:02
04 Std. Gas	0.080	10:03
05 Room Air	0.000	10:03
06 Std. Gas	0.080	10:04
07 Room Air	0.000	10:04

Lot No = 23411080A1
Cyl No = 49
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/11/2012 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.080	10:07
03 Room Air	0.000	10:08
04 Std. Gas	0.080	10:08
05 Room Air	0.000	10:09
06 Std. Gas	0.081	10:09
07 Room Air	0.000	10:09

Lot No = 23411080A1
Cyl No = 49
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/11/2012 10:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.081	10:57
03 Room Air	0.000	10:58
04 Std. Gas	0.080	10:58
05 Room Air	0.000	10:58
06 Std. Gas	0.080	10:59
07 Room Air	0.000	10:59

Lot No = 23411080A1
Cyl No = 49
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC