

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004953

Location: TOXL

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5132</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>12050</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3846</u> Lot No. <u> </u> Rep AC <u>0.100 AC + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>15</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.079</u>
Test 3 <u>0.081</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 499 psi Reg 500 psi

M. Remarks/Maintenance record: Annual inspection, replaced o-rings on breath hose and simulator return.

11 Apr 13
Date

Deb Kashner
Field Inspector's Signature

11 APR 13
Date

Charles E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:14

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNOpqrstuVwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNOpqrstuVwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 15
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 01/19/2010
Slope 691
Intercept -628068

IR Calibration Date: 06/08/2010
3um 9um

	3um	9um
0th Coef(*100):	-10343	-18245
1st Coef(*100):	249276	131289
2nd Coef(*100):	4014	1439
H2O adj(mg/l*10k):	697	655

***** Printer Test End *****


Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Sol.	0.021	09:23
03 Room Air	0.000	09:23
04 Std. Sol.	0.021	09:24
05 Room Air	0.000	09:25
06 Std. Sol.	0.021	09:25
07 Room Air	0.000	09:26

08 Sim Temp = 34.0°C

Simul Ser No = DR5132

Std Sol No = 12210

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

LOW AC ^{0.020 AC} ~~0.021 AC~~ ^{DX}

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:28
02 Std. Sol.	0.297	09:29
03 Room Air	0.000	09:29
04 Std. Sol.	0.297	09:30
05 Room Air	0.000	09:30
06 Std. Sol.	0.298	09:31
07 Room Air	0.000	09:32

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 12050

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:33

Test	AC	Time
01 Room Air	0.000	09:34
02 *Subject Test	INT*	09:34
03 Room Air	0.000	09:35

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 15
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 Ac + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:36

Test	AC	Time
01 Diagnostic	OK	09:37
02 Room Air	RFI*	09:37
03 Room Air	0.000	09:38

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 15
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI cheek

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:42
02 Std. Gas	0.080	09:43
03 Room Air	0.000	09:43
04 Std. Gas	0.079	09:43
05 Room Air	0.000	09:44
06 Std. Gas	0.081	09:44
07 Room Air	0.000	09:45

Lot No = 23411080A1
Cyl No = 15
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:46
02 Std. Gas	0.080	09:47
03 Room Air	0.000	09:47
04 Std. Gas	0.080	09:47
05 Room Air	0.000	09:48
06 Std. Gas	0.079	09:48
07 Room Air	0.000	09:49

Lot No = 23411080A1
Cyl No = 15
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = TOXL 8164.13.00 06/09
04/11/2013 09:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:50
02 Std. Gas	0.079	09:50
03 Room Air	0.000	09:51
04 Std. Gas	0.079	09:51
05 Room Air	0.000	09:52
06 Std. Gas	0.080	09:52
07 Room Air	0.000	09:52

Lot No = 23411080A1
Cyl No = 15
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000