

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004956

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u> </u> Rep AC <u>0.100AC</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>11611080A2</u>	Cyl No. <u>108</u>	Exp. Date <u>6/1/2013</u>
Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 245 psi Reg 240 psi

M. Remarks/Maintenance record: REPLACED EXTERNAL BATTERY 4.10.13;
ANNUAL INSPECTION

<u>11 APR 13</u> Date	<u>Charles E. Ed</u> Field Inspector's Signature
<u>11 Apr 13</u> Date	<u>Deb Kashur</u> Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 15:21

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 108
Standard Expiration? 06/01/2013
Oper No? 666666

Flow Cal. Date: 01/25/2010
Slope 672
Intercept -627011

IR Calibration Date: 02/15/2013
 3um 9um

0th Coef(*100): -33554 -34798
1st Coef(*100): 253701 131392
2nd Coef(*100): 3528 1568
H2O adj(mg/l*10k): 859 779

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 15:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:23
02 Std. Sol.	0.020	15:24
03 Room Air	0.000	15:25
04 Std. Sol.	0.019	15:25
05 Room Air	0.000	15:26
06 Std. Sol.	0.020	15:27
07 Room Air	0.000	15:27


08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

0.020 AC
Low AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 15:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Sol.	0.303	15:29
03 Room Air	0.000	15:30
04 Std. Sol.	0.303	15:30
05 Room Air	0.000	15:31
06 Std. Sol.	0.304	15:32
07 Room Air	0.000	15:32


08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 15:34

Test	AC	Time
01 Room Air	0.000	15:35
02 *Subject Test	INT*	15:35
03 Room Air	0.000	15:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 108
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 15:38

Test	AC	Time
01 Diagnostic	OK	15:38
02 Room Air	RFI*	15:39
03 Room Air	0.000	15:39

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPLKJMN
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 108
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 13:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.078	13:51
03 Room Air	0.000	13:52
04 Std. Gas	0.079	13:52
05 Room Air	0.000	13:53
06 Std. Gas	0.079	13:53
07 Room Air	0.000	13:53

Lot No = 11611080A2
Cyl No = 108
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.079	13:55
03 Room Air	0.000	13:55
04 Std. Gas	0.080	13:56
05 Room Air	0.000	13:56
06 Std. Gas	0.079	13:57
07 Room Air	0.000	13:57

Lot No = 11611080A2
Cyl No = 108
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004956
Location = TOXL 8164.13.00 06/09
04/11/2013 15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.080	15:04
03 Room Air	0.000	15:05
04 Std. Gas	0.079	15:05
05 Room Air	0.000	15:06
06 Std. Gas	0.079	15:06
07 Room Air	0.000	15:07

Lot No = 11611080A2
Cyl No = 108
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC