



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SPN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|----------------------------------------------------------------------------------------------------------|--------------------------------|------------------------------------|
| Chemical Test Operator Name: <u>Darcy W Aberle</u> | | |
| Location: <u>Williston LEC</u> | | |
| Intoxilyzer® Serial Number: <u>80-004189</u> | | |
| Gas Lot Number: <u>659357</u> | Gas Cylinder Number: <u>45</u> | Expiration Date: <u>09/03/2011</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed | | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) | | |
| Test 1: 0. <u>078</u> % | | |
| Test 2: 0. <u>078</u> % | | |
| Test 3: 0. <u>077</u> % | | |
| (ATTACH TEST RECORD) | | |

Darcy W Aberle
Chemical Test Operator Signature

04-12-2010
Date

Deb Keshur
Reviewed By

10 May 2010
Date

80-004189
P. 72/73 AM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WILL 8164.13.00 06/09
04/12/2010 21:43

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 21:45 |
| 02 Std. Gas | 0.078 | 21:45 |
| 03 Room Air | 0.000 | 21:46 |
| 04 Std. Gas | 0.078 | 21:46 |
| 05 Room Air | 0.000 | 21:47 |
| 06 Std. Gas | 0.077 | 21:47 |
| 07 Room Air | 0.000 | 21:48 |

Lot No = 659357
Cyl No = 45
Exp Date = 09/03/2011
County = 53 Oper No. = 010101



Operator Signature
DARCY ABERLE

Remarks:

Form 106-18000

80-004189
P. 73/73 KAL