

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006492

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5114 Lot No. 201505B Rep AC 0.030 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 1331408043 Cyl No. 028 Exp. Date 6/5/2016

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 454 psi Reg ~475 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

4.12.2016
Date

Charles E. Eden
Field Inspector's Signature

12 Apr 16
Date

Dub Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 15:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:51
02 Std. Sol.	0.027	15:52
03 Room Air	0.000	15:53
04 Std. Sol.	0.028	15:53
05 Room Air	0.000	15:54
06 Std. Sol.	0.027	15:55
07 Room Air	0.000	15:55

08 Sim Temp = 34.0°C

Simul Ser No = DR5114
Std Sol No = 201505B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC 0.030 AC*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 15:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:57
02 Std. Sol.	0.399	15:58
03 Room Air	0.000	15:58
04 Std. Sol.	0.402	15:59
05 Room Air	0.000	16:00
06 Std. Sol.	0.401	16:00
07 Room Air	0.000	16:01

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 16:03

Test	AC	Time
01 Room Air	0.000	16:04
02 *Subject Test	INT*	16:04
03 Room Air	0.000	16:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **INTERFERENT CHECK**
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 16:05

Test	AC	Time
01 Diagnostic	OK	16:06
02 Room Air	0.000	16:06
03 *Subject Test	RFI*	16:06
04 Room Air	0.000	16:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 16:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Gas	0.080	16:08
03 Room Air	0.000	16:09
04 Std. Gas	0.080	16:09
05 Room Air	0.000	16:10
06 Std. Gas	0.080	16:10
07 Room Air	0.000	16:11

Lot No = 13314080A3

Cyl No = 28

Exp Date = 06/05/2016

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

D.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 16:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:11
02 Std. Gas	0.081	16:12
03 Room Air	0.000	16:12
04 Std. Gas	0.080	16:13
05 Room Air	0.000	16:13
06 Std. Gas	0.081	16:14
07 Room Air	0.000	16:14

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.13.00 06/09
04/12/2016 16:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:15
02 Std. Gas	0.081	16:15
03 Room Air	0.000	16:16
04 Std. Gas	0.080	16:16
05 Room Air	0.000	16:17
06 Std. Gas	0.080	16:17
07 Room Air	0.000	16:18

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000