

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006504

Location: TOXL

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3374 Lot No. 14240 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3841 Lot No. 201603A Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5144 Lot No. ICS4 Solution 0.10AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 14214080A6 Cyl No. 51 Exp. Date 07/05/2016

Test 1 0.079 Test 4 0.079 Test 7 0.080

Test 2 0.079 Test 5 0.079 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 675 psi Reg 680 psi

K. Remarks/Maintenance record: Annual inspection

12 Apr 2016
Date

12 APR. 16
Date

Deb Shanaver
Field Inspector's Signature

Charles E. Eden
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 09:40

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14214080A6
Standard Cyl #? 51
Standard Expiration? 07/05/2016
Oper No? 777777

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 08/20/2015

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -14464 | -14817 |
| 1st Coef(*100): | 255017 | 131849 |
| 2nd Coef(*100): | 5533 | 1946 |
| H2O adj(mg/l*10k): | 700 | 431 |

***** Printer Test End *****


Operator Signature
DEB SHANAVER

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 10:05

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:06 |
| 02 Std. Sol. | 0.020 | 10:07 |
| 03 Room Air | 0.000 | 10:07 |
| 04 Std. Sol. | 0.020 | 10:08 |
| 05 Room Air | 0.000 | 10:08 |
| 06 Std. Sol. | 0.021 | 10:09 |
| 07 Room Air | 0.000 | 10:10 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3374
Std Sol No = 14240
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

LOW AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 10:11

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:12 |
| 02 Std. Sol. | 0.306 | 10:13 |
| 03 Room Air | 0.000 | 10:13 |
| 04 Std. Sol. | 0.306 | 10:14 |
| 05 Room Air | 0.000 | 10:14 |
| 06 Std. Sol. | 0.307 | 10:15 |
| 07 Room Air | 0.000 | 10:16 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3841
Std Sol No = 201603A
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

High AC 0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 11:59

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 11:59 |
| 02 *Subject Test | INT* | 12:00 |
| 03 Room Air | 0.000 | 12:00 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 14214080A6
Cyl No = 51
Expiration Date = 07/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Interferent Detected 0.10Ac + 0.05% acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 10:21

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 10:21 |
| 02 Room Air | RFI* | 10:21 |
| 03 Room Air | 0.000 | 10:22 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 14214080A6
Cyl No = 51
Expiration Date = 07/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 09:53

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 09:54 |
| 02 Std. Gas | 0.079 | 09:54 |
| 03 Room Air | 0.000 | 09:55 |
| 04 Std. Gas | 0.079 | 09:55 |
| 05 Room Air | 0.000 | 09:56 |
| 06 Std. Gas | 0.080 | 09:56 |
| 07 Room Air | 0.000 | 09:56 |

Lot No = 14214080A6
Cyl No = 51
Exp Date = 07/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
04/12/2016 10:00

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:01 |
| 02 Std. Gas | 0.080 | 10:01 |
| 03 Room Air | 0.000 | 10:01 |
| 04 Std. Gas | 0.080 | 10:02 |
| 05 Room Air | 0.000 | 10:02 |
| 06 Std. Gas | 0.080 | 10:03 |
| 07 Room Air | 0.000 | 10:03 |

Lot No = 14214080A6
Cyl No = 51
Exp Date = 07/05/2016
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC