

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006494

Location: TOXL

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5114 Lot No. 201505B Rep AC 0.030 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP306B Lot No. 14102 Rep AC 0.400 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 1331408043 Cyl No. 028 Exp. Date 6/5/2016

Test 1 0.080 Test 4 0.080 Test 7 0.081

Test 2 0.081 Test 5 0.081 Test 8 0.081

Test 3 0.081 Test 6 0.080 Test 9 0.082

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 537 psi Reg ~525 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

4.12.2016
Date

Charles E. Edr
Field Inspector's Signature

12 Apr 2016
Date

Dub Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:29

***** Printer Test *****

abcdefghijklmnopqrstu vwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

abcdefghijklmnopqrstu vwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 28
Standard Expiration? 06/05/2016
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 655
Intercept -624238

IR Calibration Date: 08/18/2015
3um 9um

0th Coef(*100):	-24880	-16926
1st Coef(*100):	256511	140602
2nd Coef(*100):	4127	1722
H2O adj(mg/l*10k):	783	426

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Sol.	0.028	13:35
03 Room Air	0.000	13:35
04 Std. Sol.	0.029	13:36
05 Room Air	0.000	13:36
06 Std. Sol.	0.029	13:37
07 Room Air	0.000	13:38

08 Sim Temp = 34.0°C

Simul Ser No = DR5114

Std Sol No = 201505B

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC 0.030 AC*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Sol.	0.402	13:41
03 Room Air	0.000	13:41
04 Std. Sol.	0.404	13:42
05 Room Air	0.000	13:42
06 Std. Sol.	0.406	13:43
07 Room Air	0.000	13:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 14102

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks: HIGH AC 0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:46

Test	AC	Time
01 Room Air	0.000	13:47
02 *Subject Test	INT*	13:47
03 Room Air	0.000	13:48

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:48

Test	AC	Time
01 Diagnostic	OK	13:49
02 Room Air	0.000	13:49
03 *Subject Test	RFI*	13:49
04 Room Air	0.000	13:50

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:52
02 Std. Gas	0.080	13:52
03 Room Air	0.000	13:52
04 Std. Gas	0.081	13:53
05 Room Air	0.000	13:53
06 Std. Gas	0.081	13:54
07 Room Air	0.000	13:54

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.080	13:56
03 Room Air	0.000	13:57
04 Std. Gas	0.081	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.080	13:58
07 Room Air	0.000	13:59

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006494
Location = TOXL 8164.13.00 06/09
04/12/2016 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:59
02 Std. Gas	0.081	14:00
03 Room Air	0.000	14:00
04 Std. Gas	0.081	14:00
05 Room Air	0.000	14:01
06 Std. Gas	0.082	14:01
07 Room Air	0.000	14:02

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000