

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006498

Location: TOXL

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5114 Lot No. 201505B Rep AC 0.030 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10% AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A3 Cyl No. 028 Exp. Date 6/5/2016

Test 1 0.081 Test 4 0.081 Test 7 0.082

Test 2 0.081 Test 5 0.081 Test 8 0.081

Test 3 0.081 Test 6 0.080 Test 9 0.081

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 505 psi Reg ~500 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

4.12.2016
Date

Charles E. Edler
Field Inspector's Signature

12 Apr 2016
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:17

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 28
Standard Expiration? 06/05/2016
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 669
Intercept -637469

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100): -28161 -22508
1st Coef(*100): 286657 143843
2nd Coef(*100): 2341 1281
H2O adj (mg/l*10k): 776 451

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:21
02 Std. Sol.	0.028	14:22
03 Room Air	0.000	14:22
04 Std. Sol.	0.028	14:23
05 Room Air	0.000	14:24
06 Std. Sol.	0.028	14:24
07 Room Air	0.000	14:25

08 Sim Temp = 34.0°C

Simul Ser No = DR5114

Std Sol No = 201505B

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC 0.030 AC*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Sol.	0.399	14:27
03 Room Air	0.000	14:28
04 Std. Sol.	0.400	14:29
05 Room Air	0.000	14:29
06 Std. Sol.	0.400	14:30
07 Room Air	0.000	14:30

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC 0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:32

Test	AC	Time
01 Room Air	0.000	14:33
02 *Subject Test	INT*	14:34
03 Room Air	0.000	14:34

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:36

Test	AC	Time
01 Diagnostic	OK	14:37
02 Room Air	0.000	14:37
03 *Subject Test	RFI*	14:37
04 Room Air	0.000	14:38

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:41
02 Std. Gas	0.081	14:41
03 Room Air	0.000	14:42
04 Std. Gas	0.081	14:42
05 Room Air	0.000	14:43
06 Std. Gas	0.081	14:43
07 Room Air	0.000	14:44

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.081	14:45
03 Room Air	0.000	14:45
04 Std. Gas	0.081	14:46
05 Room Air	0.000	14:46
06 Std. Gas	0.080	14:47
07 Room Air	0.000	14:47

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = TOXL 8164.13.00 06/09
04/12/2016 14:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.082	14:49
03 Room Air	0.000	14:50
04 Std. Gas	0.081	14:50
05 Room Air	0.000	14:50
06 Std. Gas	0.081	14:51
07 Room Air	0.000	14:51

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AR