

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006500

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5114 Lot No. 201505B Rep AC 0.030
201505B
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 14102 Rep AC 0.400
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC+0.05% ACETONE
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 1331408043 Cyl No. 028 Exp. Date 6/5/2016

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 610 psi Reg -600 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

4.12.2016
Date

Charles E. Edler
Field Inspector's Signature

12 Apr 2016
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 11:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 25
Standard Expiration? 12/05/2016
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 681
Intercept -680985

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100): -3148 -17259
1st Coef(*100): 254475 138520
2nd Coef(*100): 1954 1329
H2O adj(mg/l*10k): 193 414

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 11:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Sol.	0.027	11:19
03 Room Air	0.000	11:19
04 Std. Sol.	0.028	11:20
05 Room Air	0.000	11:21
06 Std. Sol.	0.029	11:21
07 Room Air	0.000	11:22

08 Sim Temp = 34.0°C

Simul Ser No = DR5114

Std Sol No = 201505B

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC 0.030 AC*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 11:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:24
02 Std. Sol.	0.399	11:25
03 Room Air	0.000	11:26
04 Std. Sol.	0.401	11:26
05 Room Air	0.000	11:27
06 Std. Sol.	0.401	11:28
07 Room Air	0.000	11:28

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC 0.400 AC*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 12:24

Test	AC	Time
01 Room Air	0.000	12:24
02 *Subject Test	INT*	12:25
03 Room Air	0.000	12:26

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 12:26

Test	AC	Time
01 Diagnostic	OK	12:27
02 Room Air	0.000	12:27
03 *Subject Test	RFI*	12:27
04 Room Air	0.000	12:28

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 12:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:30
02 Std. Gas	0.080	12:31
03 Room Air	0.000	12:31
04 Std. Gas	0.080	12:32
05 Room Air	0.000	12:32
06 Std. Gas	0.081	12:32
07 Room Air	0.000	12:33

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 12:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:34
02 Std. Gas	0.081	12:34
03 Room Air	0.000	12:35
04 Std. Gas	0.081	12:35
05 Room Air	0.000	12:36
06 Std. Gas	0.081	12:36
07 Room Air	0.000	12:36

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = TOXL 8164.13.00 06/09
04/12/2016 12:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:37
02 Std. Gas	0.081	12:38
03 Room Air	0.000	12:38
04 Std. Gas	0.081	12:38
05 Room Air	0.000	12:39
06 Std. Gas	0.081	12:39
07 Room Air	0.000	12:40

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Form 106-I8000