

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004952

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5134 Lot No. Rep AC 0.294 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. Rep AC 0.10 + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 05911080A1 Cyl No. 18 Exp. Date 02/01/2013
Test 1 0.079 Test 4 0.079 Test 7 0.079
Test 2 0.080 Test 5 0.080 Test 8 0.079
Test 3 0.080 Test 6 0.080 Test 9 0.080
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 826 psi Reg 820 psi

M. Remarks/Maintenance record: Annual inspection

13 Apr. 2011
Date
[Signature]
Date ²² Reviewed by

Deb Kashur
Field Inspector's Signature
4-13-11
Reviewed by ²² Date

80-004952
P. 43 of 51
[Signature]

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = LISB 8164.13.00 06/09
04/13/2011 09:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05911080A1
Standard Cyl #? 18
Standard Expiration? 02/01/2013
Oper No? 777777

Flow Cal. Date: 01/19/2010
Slope 694
Intercept -638616

IR Calibration Date: 01/19/2010
 3um 9um


0th Coef(*100): -32832 -34603
1st Coef(*100): 262566 131011
2nd Coef(*100): 1418 975
H2O adj(mg/l*10k): 686 550

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 09:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:47
02 Std. Sol.	0.018	09:50
03 Room Air	0.000	09:50
04 Std. Sol.	0.018	09:51
05 Room Air	0.000	09:51
06 Std. Sol.	0.019	09:52
07 Room Air	0.000	09:53

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

low AC 0.021 AC

Form 106-I8000

80-004952
p. 45 of 51
SM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 10:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:11
02 Std. Sol.	0.289	10:12
03 Room Air	0.000	10:13
04 Std. Sol.	0.289	10:14
05 Room Air	0.000	10:14
06 Std. Sol.	0.290	10:15
07 Room Air	0.000	10:16

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.294 AC

Form 106-I8000

80-004952
P. 46 of 51
JAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 11:09

Test	AC	Time
01 Room Air	0.000	11:09
02 *Subject Test	INT*	11:10
03 Room Air	0.000	11:11

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 05911080A1
Cyl No = 18
Expiration Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10 AC + acetone

Form 106-I8000

80-004952
P. 47 of 51
JKM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 11:05

Test	AC	Time
01 Diagnostic	OK	11:06
02 Room Air	RFI*	11:06
03 Room Air	0.000	11:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 05911080A1
Cyl No = 18
Expiration Date = 02/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004952
p. 48 of 51
SM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 10:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:49
02 Std. Gas	0.079	10:49
03 Room Air	0.000	10:50
04 Std. Gas	0.080	10:50
05 Room Air	0.000	10:51
06 Std. Gas	0.080	10:51
07 Room Air	0.000	10:51

Lot No = 05911080A1
Cyl No = 18
Exp Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Form 106-I8000

*80-004952
p. 49 of 51
JKM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 10:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:55
02 Std. Gas	0.079	10:55
03 Room Air	0.000	10:56
04 Std. Gas	0.080	10:56
05 Room Air	0.000	10:57
06 Std. Gas	0.080	10:57
07 Room Air	0.000	10:58

Lot No = 05911080A1
Cyl No = 18
Exp Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

*80-004952
P. 50 of 51
DM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.13.00 06/09
04/13/2011 10:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:58
02 Std. Gas	0.079	10:59
03 Room Air	0.000	10:59
04 Std. Gas	0.079	11:00
05 Room Air	0.000	11:00
06 Std. Gas	0.080	11:01
07 Room Air	0.000	11:01

Lot No = 05911080A1
Cyl No = 18
Exp Date = 02/01/2013
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004952
P. 51 of 51
JEM