



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Richard D. Johnson

Location: McLean LEC Washburn ND

Intoxilyzer Serial No.: 80-004196

Lot Number: 659357

Cylinder Number: 47

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.081 %

(ATTACH TEST RECORD)

Test 3: 0.082 %

Richard D. Johnson 4-14-10  
 Chemical Test Operator Signature

Date Cylinder Changed

Deb Keshur  
 Reviewed By

19 Apr 2010  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-004196  
 76/77 PM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
04/14/2010      14:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:32
02 Std. Gas	0.081	14:33
03 Room Air	0.000	14:33
04 Std. Gas	0.081	14:33
05 Room Air	0.000	14:34
06 Std. Gas	0.082	14:34
07 Room Air	0.000	14:35

Lot No = 659357  
Cyl No = 47  
Exp Date = 09/03/2011  
County = 28      Oper No. = 118707

  
Operator Signature  
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196  
7/77 KM