



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	<i>Berry Vannatta</i>		
Location:	<i>Moh L.</i>		
Intoxilyzer® Serial Number:	<i>80-004947</i>		
Gas Lot Number:	Gas Cylinder Number:	Expiration Date:	
<i>659358</i>	<i>27</i>	<i>8-3-11</i>	
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1: 0. <u><i>081</i></u> %			
Test 2: 0. <u><i>080</i></u> %			
Test 3: 0. <u><i>080</i></u> %			
(ATTACH TEST RECORD)			

Berry Vannatta
Chemical Test Operator Signature

4-14-11
Date

Deb Kasher
Reviewed By

19 Apr 2011
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = MOHL 8164.13.00 06/09
04/14/2011 17:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:12
02 Std. Gas	0.081	17:13
03 Room Air	0.000	17:13
04 Std. Gas	0.080	17:13
05 Room Air	0.000	17:14
06 Std. Gas	0.080	17:14
07 Room Air	0.000	17:15

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909


Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004947
p. 71 of 71
KM