

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004206

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 13070 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3374 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. 1CS2 Rep AC 0.100 AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>16512080A2</u>	Cyl No. <u>22</u>	Exp. Date <u>08/01/2014</u>
Test 1 <u>0.082</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.082</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

J. Tank pressure, Level 3 Func D Sub G P: 565 psi Reg 560 psi

K. Remarks/Maintenance record: Annual inspection, replaced simulator return o-ring.

10 Apr 2014
Date

14 APR 14 ✓
Date

Deb Kashner
Field Inspector's Signature

[Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = TOXL 8164.13.00 06/09
04/10/2014 09:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 22
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 06/10/2009
Slope 690
Intercept -645883

IR Calibration Date: 06/07/2010
3um 9um

0th Coef(*100): -17579 -21810
1st Coef(*100): 254057 127149
2nd Coef(*100): 4591 1650
H2O adj(mg/l*10k): 756 795

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = ~~WAHP-TOXL~~ *DK* 8164.13.00 06/09
04/10/2014 08:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:52
02 Std. Sol.	0.020	08:53
03 Room Air	0.000	08:53
04 Std. Sol.	0.020	08:54
05 Room Air	0.000	08:55
06 Std. Sol.	0.021	08:55
07 Room Air	0.000	08:56

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 13070

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = ~~WAHP~~ ^{TOXL} 8164.13.00 06/09
04/10/2014 DK 08:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:47
02 Std. Sol.	0.301	08:47
03 Room Air	0.000	08:48
04 Std. Sol.	0.301	08:49
05 Room Air	0.000	08:49
06 Std. Sol.	0.300	08:50
07 Room Air	0.000	08:51

08 Sim Temp = 34.0°C

Simul Ser No = DR3374
Std Sol No = 13080
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = TOXL 8164.13.00 06/09
04/10/2014 09:01

Test	AC	Time
01 Room Air	0.000	09:01
02 *Subject Test	INT*	09:02
03 Room Air	0.000	09:02

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = TOXL 8164.13.00 06/09
04/10/2014 09:03

Test	AC	Time
01 Diagnostic	OK	09:04
02 Room Air	RFI*	09:04
03 Room Air	0.000	09:05

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = ~~WAHP TOXL~~ 8164.13.00 06/09
04/10/2014 *DK* 08:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:41
02 Std. Gas	0.082	08:41
03 Room Air	0.000	08:42
04 Std. Gas	0.082	08:42
05 Room Air	0.000	08:42
06 Std. Gas	0.082	08:43
07 Room Air	0.000	08:43

Lot No = 16512080A2
Cyl No = 22
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = TOXL 8164.13.00 06/09
04/10/2014 09:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:06
02 Std. Gas	0.081	09:06
03 Room Air	0.000	09:07
04 Std. Gas	0.081	09:07
05 Room Air	0.000	09:08
06 Std. Gas	0.081	09:08
07 Room Air	0.000	09:08

Lot No = 16512080A2
Cyl No = 22
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.081 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = TOXL 8164.13.00 06/09
04/10/2014 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.081	10:07
03 Room Air	0.000	10:08
04 Std. Gas	0.082	10:08
05 Room Air	0.000	10:09
06 Std. Gas	0.081	10:09
07 Room Air	0.000	10:10

Lot No = 16512080A2
Cyl No = 22
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080Ae