

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005363

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR3803 Lot No. 13070 Rep AC 0.020 ✓
  - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR3374 Lot No. 13080 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR3375 Lot No. 1CS2 Rep AC 0.100 Ac + 0.05% acetone ✓
  - H. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 16512080A2 Cyl No. 22 Exp. Date 08/01/2014  
Test 1 0.081 Test 4 0.080 Test 7 0.081  
Test 2 0.080 Test 5 0.080 Test 8 0.080  
Test 3 0.080 Test 6 0.080 Test 9 0.080  
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 510 psi Reg 520 psi

K. Remarks/Maintenance record: Annual inspection

10 Apr 2014  
Date

14 APR 14 ✓  
Date

Deb Kashner  
Field Inspector's Signature

Charles E. Eden  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005363  
Location = TOXL                    8164.13.00 06/09  
04/10/2014                            11:36

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            16512080A2  
Standard Cyl #?            22  
Standard Expiration?       08/01/2014  
Oper No?                    777777

Flow Cal. Date:            08/19/2011  
Slope                        660  
Intercept                   -560671

IR Calibration Date:        07/24/2013

	3um	9um
0th Coef(*100):	-10816	-21290
1st Coef(*100):	272266	140951
2nd Coef(*100):	3523	1334
H2O adj(mg/l*10k):	587	662

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = TOXL      8164.13.00 06/09  
04/10/2014      11:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:09
02 Std. Sol.	0.020	11:25
03 Room Air	0.000	11:26
04 Std. Sol.	0.021	11:27
05 Room Air	0.000	11:27
06 Std. Sol.	0.020	11:28
07 Room Air	0.000	11:28

08 Sim Temp = 34.0°C

Simul Ser No = DR3803  
Std Sol No = 13070  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Low Ac      0.021 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = TOXL      8164.13.00 06/09  
04/10/2014      11:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Sol.	0.306	11:32
03 Room Air	0.000	11:32
04 Std. Sol.	0.305	11:33
05 Room Air	0.000	11:33
06 Std. Sol.	0.303	11:34
07 Room Air	0.000	11:35

08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 13080

County = 08

Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

High AC      0.300 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005363  
Location = TOXL                    8164.13.00 06/09  
04/10/2014                            11:46

Test	AC	Time
01 Diagnostic	OK	11:46
02 Room Air	RFI*	11:47
03 Room Air	0.000	11:47

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 16512080A2  
Cyl No = 22  
Expiration Date = 08/01/2014  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = TOXL      8164.13.00 06/09  
04/10/2014      11:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:03
02 Std. Gas	0.081	11:03
03 Room Air	0.000	11:04
04 Std. Gas	0.080	11:04
05 Room Air	0.000	11:04
06 Std. Gas	0.080	11:05
07 Room Air	0.000	11:05

Lot No = 16512080A2  
Cyl No = 22  
Exp Date = 08/01/2014  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = TOXL      8164.13.00 06/09  
04/10/2014      11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Gas	0.080	11:56
03 Room Air	0.000	11:57
04 Std. Gas	0.080	11:57
05 Room Air	0.000	11:58
06 Std. Gas	0.080	11:58
07 Room Air	0.000	11:59

Lot No = 16512080A2  
Cyl No = 22  
Exp Date = 08/01/2014  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = TOXL      8164.13.00 06/09  
04/10/2014      11:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:00
02 Std. Gas	0.081	12:00
03 Room Air	0.000	12:01
04 Std. Gas	0.080	12:01
05 Room Air	0.000	12:01
06 Std. Gas	0.080	12:02
07 Room Air	0.000	12:02

Lot No = 16512080A2  
Cyl No = 22  
Exp Date = 08/01/2014  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*