



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number <u>60-005955</u>	Instrument Location <u>MANNING - DCSO</u>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>FIELD OPERATOR CHANGE</u>	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): NONE
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <u>04/13/16</u>
Reviewed By (Crime Laboratory Use Only) <u>Deb Shanaver</u>	Date <u>14 Apr 16</u>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.13.00 06/09
04/13/2016 06:46

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 29
Standard Expiration? 09/05/2017
Oper No? 130635

Flow Cal. Date: 06/21/2013
Slope 757
Intercept -837385

IR Calibration Date: 06/21/2013
 3um 9um

0th Coef(*100): -19897 -22878
1st Coef(*100): 268381 131979
2nd Coef(*100): 5157 1473
H2O adj(mg/l*10k): 717 600

***** Printer Test End *****



Operator Signature
HUNTER RHOADES

Remarks: **INSTALL**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.13.00 06/09
04/13/2016 06:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	06:47
02 Std. Gas	0.083	06:48
03 Room Air	0.000	06:48
04 Std. Gas	0.084	06:49
05 Room Air	0.000	06:49
06 Std. Gas	0.084	06:50
07 Room Air	0.000	06:50

Lot No = 19615080A5
Cyl No = 29
Exp Date = 09/05/2017
County = 13 Oper No. = 130635



Operator Signature
HUNTER RHOADES

Remarks: *ACA TEST*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.13.00 06/09
04/13/2016 06:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	06:51
02 Std. Gas	0.083	06:52
03 Room Air	RFI*	06:52
04 Room Air	0.000	06:52

*Invalid Test
Inhibited - RFI

Lot No = 19615080A5
Cyl No = 29
Exp Date = 09/05/2017
County = 13 Oper No. = 130635


Operator Signature
HUNTER RHOADES

Remarks: RFI TEST