

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-002669

Location: TOXL

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3066 Lot No. 15160 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC+0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A³CCE Cyl No. 028 Exp. Date 6/5/2016

Test 1 0.079 Test 4 0.078 Test 7 0.078

Test 2 0.078 Test 5 0.078 Test 8 0.078

Test 3 0.078 Test 6 0.078 Test 9 0.078

Average 0.078

J. Tank pressure, Level 3 Func D Sub G P: 143 psi Reg ~150 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, REPLACED SIMULATOR RETURN O-RING

4.14.2016
Date

Charles E. Elden
Field Inspector's Signature

14 Apr 16
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:10

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=-|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=-|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 28
Standard Expiration? 06/05/2016
Oper No? 666666

Flow Cal. Date: 09/09/2011
Slope 662
Intercept -510765

IR Calibration Date: 02/06/2015

	3um	9um
0th Coef (*100):	-31426	-24477
1st Coef (*100):	275748	139415
2nd Coef (*100):	1078	912
H2O adj (mg/l*10k):	429	271

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:11
02 Std. Sol.	0.017	15:12
03 Room Air	0.000	15:13
04 Std. Sol.	0.017	15:14
05 Room Air	0.000	15:14
06 Std. Sol.	0.017	15:15
07 Room Air	0.000	15:15

08 Sim Temp = 34.0°C

Simul Ser No = MP3066
Std Sol No = 15160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:17
02 Std. Sol.	0.397	15:18
03 Room Air	0.000	15:18
04 Std. Sol.	0.400	15:19
05 Room Air	0.000	15:20
06 Std. Sol.	0.399	15:20
07 Room Air	0.000	15:21

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:23

Test	AC	Time
01 Room Air	0.000	15:24
02 *Subject Test	INT*	15:24
03 Room Air	0.000	15:25

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:25

Test	AC	Time
01 Diagnostic	OK	15:26
02 Room Air	0.000	15:26
03 *Subject Test	RFI*	15:26
04 Room Air	0.000	15:27

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:27
02 Std. Gas	0.079	15:28
03 Room Air	0.000	15:28
04 Std. Gas	0.078	15:29
05 Room Air	0.000	15:29
06 Std. Gas	0.078	15:29
07 Room Air	0.000	15:30

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:31
02 Std. Gas	0.078	15:31
03 Room Air	0.000	15:32
04 Std. Gas	0.078	15:32
05 Room Air	0.000	15:32
06 Std. Gas	0.078	15:33
07 Room Air	0.000	15:33

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.13.00 06/09
04/14/2016 15:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:34
02 Std. Gas	0.078	15:35
03 Room Air	0.000	15:35
04 Std. Gas	0.078	15:35
05 Room Air	0.000	15:36
06 Std. Gas	0.078	15:36
07 Room Air	0.000	15:37

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC