



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Shannon Liebold				
Location:	Mercer County Sheriff's Dept.				
Intoxilyzer® Serial Number:	80-003060				
Gas Lot Number:	659357	Gas Cylinder Number:	43	Expiration Date:	09/03/2011
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.081 %					
Test 2: 0.079 %					
Test 3: 0.081 %					
(ATTACH TEST RECORD)					

Shannon Liebold
Chemical Test Operator Signature

4-15-10
Date

Deb Kashner
Reviewed By

19 Apr 2010
Date

80-003060
7/77 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
04/15/2010 11:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:04
02 Std. Gas	0.081	11:04
03 Room Air	0.000	11:05
04 Std. Gas	0.079	11:05
05 Room Air	0.000	11:06
06 Std. Gas	0.081	11:06
07 Room Air	0.000	11:07

Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060
7/77 KM