

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005953

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5115</u> Lot No. <u>12111</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5143</u> Lot No. <u>13080</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5135</u> Lot No. <u>ICS#2</u> Rep AC <u>0.100AC+ 0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>16512080A2</u>	Cyl No. <u>034</u>	Exp. Date <u>08/01/2014</u>
Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.083</u>
Test 2 <u>0.081</u>	Test 5 <u>0.083</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.083</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

J. Tank pressure, Level 3 Func D Sub G P: 118 psi Reg 125 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

15 APR 2014
Date

Charles E. Ed
Field Inspector's Signature

15 Apr 2014
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 09:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 34
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 674
Intercept -679248

IR Calibration Date: 06/21/2013

	3um	9um
0th Coef(*100):	-14815	-18665
1st Coef(*100):	263530	132860
2nd Coef(*100):	3000	1324
H2O adj(mg/l*10k):	578	483

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 09:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:35
02 Std. Sol.	0.020	09:36
03 Room Air	0.000	09:37
04 Std. Sol.	0.021	09:37
05 Room Air	0.000	09:38
06 Std. Sol.	0.021	09:39
07 Room Air	0.000	09:39

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

0.020 AC

Low AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 09:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:41
02 Std. Sol.	0.298	09:42
03 Room Air	0.000	09:43
04 Std. Sol.	0.299	09:44
05 Room Air	0.000	09:44
06 Std. Sol.	0.301	09:45
07 Room Air	0.000	09:45

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 13080
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 09:48

Test	AC	Time
01 Room Air	0.000	09:49
02 *Subject Test	INT*	09:49
03 Room Air	0.000	09:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 34
Expiration Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK
0.100AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 09:53

Test	AC	Time
01 Diagnostic	OK	09:54
02 Room Air	RFI*	09:54
03 Room Air	0.000	09:54

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 34
Expiration Date = 08/01/2014
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 09:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:55
02 Std. Gas	0.082	09:55
03 Room Air	0.000	09:56
04 Std. Gas	0.081	09:56
05 Room Air	0.000	09:57
06 Std. Gas	0.082	09:57
07 Room Air	0.000	09:58

Lot No = 16512080A2
Cyl No = 34
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

0.080 AC

CALIBRATION CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 10:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:00
02 Std. Gas	0.082	10:01
03 Room Air	0.000	10:01
04 Std. Gas	0.083	10:02
05 Room Air	0.000	10:02
06 Std. Gas	0.083	10:03
07 Room Air	0.000	10:03

Lot No = 16512080A2
Cyl No = 34
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.080 AC

CALIBRATION CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = TOXL 8164.13.00 06/09
04/15/2014 10:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:04
02 Std. Gas	0.083	10:05
03 Room Air	0.000	10:05
04 Std. Gas	0.082	10:05
05 Room Air	0.000	10:06
06 Std. Gas	0.082	10:06
07 Room Air	0.000	10:07

Lot No = 16512080A2
Cyl No = 34
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.080AC

CALIBRATION CHECK