



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Hinton, Rick W.

Location: Cass County Jail

Intoxilyzer Serial No.: 80-004205

Lot Number: 659

Cylinder Number: 40

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082%

Test 2: 0.082%

Test 3: 0.082%

(ATTACH TEST RECORD)

[Signature]
 Chemical Test Operator Signature

NA
 Date Cylinder Changed

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004205
 P. 1/2
 P. 68/69


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
04/16/2010 08:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:36
02 Std. Gas	0.082	08:36
03 Room Air	0.000	08:37
04 Std. Gas	0.082	08:37
05 Room Air	0.000	08:38
06 Std. Gas	0.082	08:38
07 Room Air	0.000	08:39

Lot No = 659357
Cyl No = 40
Exp Date = 09/03/2011
County = 09 Oper No. = 029920



Operator Signature
RICK HINTON

Remarks:

Form 106-I8000

80-004205
P. 69/69
/09