



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>ZLOTKOWSKI, GERALD R.</u>		
Location: <u>ASHL</u>		
Intoxilyzer® Serial Number: <u>80-004945</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>32</u>	Expiration Date: <u>12/23/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed <u>NEW INSTALLATION OF 8000</u>		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.080</u> % Test 2: <u>0.080</u> % Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

*Gerald R. Zlotkowski* M.C.S.O.  
Chemical Test Operator Signature

4/16/2010  
Date

*Deb Kadner*  
Reviewed By

07 June 2010  
Date


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = ASHL      8164.13.00 06/09  
04/16/2010      16:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:45
02 Std. Gas	0.080	16:46
03 Room Air	0.000	16:46
04 Std. Gas	0.080	16:47
05 Room Air	0.000	16:47
06 Std. Gas	0.080	16:47
07 Room Air	0.000	16:48

Lot No = 709352  
Cyl No = 32  
Exp Date = 12/23/2011  
County = 26      Oper No. = 100631

  
\_\_\_\_\_  
Operator Signature  
GERARD R. ZLOTKOWSKI

Remarks:

Form 106-I8000

80-004945  
p. 33 of 33  
KM