

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004954

Location: TOXL

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>    </u> Rep AC <u>0.019</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5187</u> Lot No. <u>    </u> Rep AC <u>0.285</u>                               | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>G8766</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>709352</u>	Cyl No. <u>1</u>	Exp. Date <u>23 Dec 2011</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.081</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 581 psi Reg 580 psi

M. Remarks/Maintenance record: Annual inspection

16 Apr 2010  
Date

Deb Kashur  
Field Inspector's Signature

16 Apr 2010  
Date

Maryn Pearson  
Reviewed by

80-004954  
32/40 KM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
04/16/2010      11:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:08
02 Std. Sol.	0.017	11:09
03 Room Air	0.000	11:09
04 Std. Sol.	0.018	11:10
05 Room Air	0.000	11:11
06 Std. Sol.	0.019	11:11
07 Room Air	0.000	11:12

08 Sim Temp = 34.0°C

Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.019 AC*

Form 106-I8000

*80-004954  
3/4/10 KM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
04/16/2010      10:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:41
02 Std. Sol.	0.294	10:42
03 Room Air	0.000	10:43
04 Std. Sol.	0.294	10:43
05 Room Air	0.000	10:44
06 Std. Sol.	0.292	10:45
07 Room Air	0.000	10:45

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.285 AC*

Form 106-I8000

80-004954  
35/40 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004954  
Location = TOXL                8164.13.00 06/09  
04/16/2010                        11:14

Test	AC	Time
01 Room Air	0.000	11:15
02 *Subject Test	INT*	11:16
03 Room Air	0.000	11:16

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08                        Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

80-004954  
3/2/10 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004954  
Location = TOXL                8164.13.00 06/09  
04/16/2010                        09:16

Test	AC	Time
01 Diagnostic	OK	09:17
02 Room Air	RFI*	09:17
03 Room Air	0.000	09:18

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 1  
Expiration Date = 12/23/2011  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004954  
37/10 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004954  
Location = TOXL                8164.13.00 06/09  
04/16/2010                        09:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:20
02 Std. Gas	0.080	09:20
03 Room Air	0.000	09:21
04 Std. Gas	0.080	09:21
05 Room Air	0.000	09:22
06 Std. Gas	0.081	09:22
07 Room Air	0.000	09:22

Lot No = 709352  
Cyl No = 1  
Exp Date = 12/23/2011  
County = 08                      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004954  
3/10 FH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
04/16/2010      09:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Gas	0.081	09:28
03 Room Air	0.000	09:28
04 Std. Gas	0.080	09:29
05 Room Air	0.000	09:29
06 Std. Gas	0.080	09:29
07 Room Air	0.000	09:30

Lot No = 709352  
Cyl No = 1  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004954  
3/10 *KTM*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
04/16/2010      09:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:32
02 Std. Gas	0.080	09:32
03 Room Air	0.000	09:33
04 Std. Gas	0.080	09:33
05 Room Air	0.000	09:33
06 Std. Gas	0.080	09:34
07 Room Air	0.000	09:34

Lot No = 709352  
Cyl No = 1  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

*80-004954  
4/16/10 KM*