

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004204

Location: TOX LAB

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5190</u> Lot No. <u>---</u> Rep AC <u>0.019 AC</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5113</u> Lot No. <u>---</u> Rep AC <u>0.285 AC</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5135</u> Lot No. <u>---</u> Rep AC <u>0.10 AC + 0.05% Acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>709352</u>	Cyl No. <u>41</u>	Exp. Date <u>10/23/2011</u>
Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 851 psi Reg 850 psi

M. Remarks/Maintenance record: Annual Inspection

<u>16 Apr 2010</u> Date	<u>Mary Pearson</u> Field Inspector's Signature
<u>19 Apr 2010</u> Date	<u>Deb Kashur</u> Reviewed by

80-004204  
38/42 12/20

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/16/2010                      14:43

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            709352  
Standard Cyl #?            41  
Standard Expiration?       12/23/2011  
Oper No?                    107501

Flow Cal. Date:            06/04/2009  
Slope                      702  
Intercept                  -590070

IR Calibration Date:       01/13/2010  
                                    3um                      9um

-----  
0th Coef(\*100):            -19674                      -24174  
1st Coef(\*100):            258610                      128392  
2nd Coef(\*100):            659                         701  
H2O adj(mg/l\*10k):       544                         551

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
MARGY PEARSON

Remarks:

Print

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
04/16/2010      14:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:46
02 Std. Sol.	0.017	14:49
03 Room Air	0.000	14:49
04 Std. Sol.	0.017	14:50
05 Room Air	0.000	14:50
06 Std. Sol.	0.017	14:51
07 Room Air	0.000	14:52

08 Sim Temp = 34.0°C

Simul Ser No = 00000000  
Std Sol No = 000000  
County = 08      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Low AC 0.019 AC*

Form 106-I8000

*80-004204  
4/16/10*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
04/16/2010      14:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:55
02 Std. Sol.	0.277	14:56
03 Room Air	0.000	14:57
04 Std. Sol.	0.275	14:58
05 Room Air	0.000	14:58
06 Std. Sol.	0.275	14:59
07 Room Air	0.000	14:59

08 Sim Temp = 34.0°C

Simul Ser No = 00000000  
Std Sol No = 000000  
County = 08      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Linearity 0.285AC*

Form 106-I8000

*80-004204  
4/16/10*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/16/2010                      15:00

Test	AC	Time
01 Room Air	0.000	15:03
02 *Subject Test	INT*	15:03
03 Room Air	0.000	15:04

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 00000000  
Std Sol No = 000000  
County = 08                      Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

0.10 AC + 0.05% Acetone

Form 106-I8000

80-004204  
4/16/10 KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
04/16/2010                            15:04

Test	AC	Time
01 Diagnostic	OK	15:05
02 Room Air	RFI*	15:05
03 Room Air	0.000	15:06

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 00000000  
Std Sol No = 000000  
County = 08                            Oper No. = 107501

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
MARGY PEARSON

Remarks:

*RFI*

Form 106-I8000

80-004204  
4/3/44 KM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
04/16/2010                            14:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:29
02 Std. Gas	0.078	14:30
03 Room Air	0.000	14:30
04 Std. Gas	0.079	14:31
05 Room Air	0.000	14:31
06 Std. Gas	0.079	14:32
07 Room Air	0.000	14:32

Lot No = 709352  
Cyl No = 41  
Exp Date = 12/23/2011  
County = 08                            Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check*

Form 106-I8000

*80-004204  
4/4/16 PM*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
04/16/2010                        14:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:34
02 Std. Gas	0.079	14:34
03 Room Air	0.000	14:34
04 Std. Gas	0.078	14:35
05 Room Air	0.000	14:35
06 Std. Gas	0.079	14:36
07 Room Air	0.000	14:36

Lot No = 709352  
Cyl No = 41  
Exp Date = 12/23/2011  
County = 08                      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check*

Form 106-I8000

*80-0084204  
4/5/16 JRM*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
04/16/2010                            14:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:38
02 Std. Gas	0.079	14:38
03 Room Air	0.000	14:39
04 Std. Gas	0.079	14:39
05 Room Air	0.000	14:40
06 Std. Gas	0.079	14:40
07 Room Air	0.000	14:41

Lot No = 709352  
Cyl No = 41  
Exp Date = 12/23/2011  
County = 08                            Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check*

Form 106-I8000

*80-004204  
2/6/16 YAN*