

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004938

Location: TOXL

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u> </u> Rep AC <u> </u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5187</u> Lot No. <u> </u> Rep AC <u> </u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>G-8766</u> Lot No. <u> </u> Rep AC <u> </u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>709352</u>	Cyl No. <u>21</u>	Exp. Date <u>23 Dec 2010</u> ^{2011 DK}
Test 1 <u>0.078</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.078</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 381 psi Reg 390 psi

M. Remarks/Maintenance record: Annual inspection

16 Apr 2010
Date

Deb Kashur
Field Inspector's Signature

16 Apr 2010
Date

Mary Plauson
Reviewed by

80-004938
34/42 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 09:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 21
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 01/25/2010
Slope 712
Intercept -604165

IR Calibration Date: 02/05/2010

	3um	9um
0th Coef(*100):	-32364	-29837
1st Coef(*100):	263210	133607
2nd Coef(*100):	406	576
H2O adj(mg/l*10k):	864	679

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 11:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:23
02 Std. Sol.	0.019	11:25
03 Room Air	0.000	11:25
04 Std. Sol.	0.019	11:26
05 Room Air	0.000	11:26
06 Std. Sol.	0.019	11:27
07 Room Air	0.000	11:27

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-18000

80-004938
3/12 1st

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 11:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:06
02 Std. Sol.	0.288	11:07
03 Room Air	0.000	11:07
04 Std. Sol.	0.287	11:08
05 Room Air	0.000	11:08
06 Std. Sol.	0.288	11:09
07 Room Air	0.000	11:10

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285 AC

Form 106-I8000

80-004938
37/42 KA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 11:11

Test	AC	Time
01 Room Air	0.000	11:12
02 *Subject Test	INT*	11:12
03 Room Air	0.000	11:13

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Interferent check 0.100 AC + acetone

Form 106-I8000

80-004938
3/8/12 KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 10:54

Test	AC	Time
01 Diagnostic	OK	10:55
02 Room Air	RFI*	10:55
03 Room Air	0.000	10:55

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 21
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004938
39/42 KA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 09:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Gas	0.078	09:24
03 Room Air	0.000	09:25
04 Std. Gas	0.078	09:25
05 Room Air	0.000	09:26
06 Std. Gas	0.078	09:26
07 Room Air	0.000	09:27

Lot No = 709352
Cyl No = 21
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

*80-004938
4/16/12 JK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 09:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:30
02 Std. Gas	0.078	09:31
03 Room Air	0.000	09:31
04 Std. Gas	0.079	09:31
05 Room Air	0.000	09:32
06 Std. Gas	0.078	09:32
07 Room Air	0.000	09:33

Lot No = 709352
Cyl No = 21
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check ^{JK} ~~0.00~~ 0.080 AC

Form 106-I8000

80-004938
4/1/12 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/16/2010 09:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:34
02 Std. Gas	0.078	09:34
03 Room Air	0.000	09:35
04 Std. Gas	0.079	09:35
05 Room Air	0.000	09:36
06 Std. Gas	0.078	09:36
07 Room Air	0.000	09:36

Lot No = 709352
Cyl No = 21
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004938
4/2/12 KM