

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005952

Location: ~~TOX~~^{DK} BCSO

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.) ICS 2 ✓
Sim SN DR3846 Lot No. ICS 2 OK Rep AC 0.10 AC + 0.05% acetone
- H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 69 Exp. Date 08/01/2014

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 942 psi Reg 960 psi

K. Remarks/Maintenance record: Annual inspection

16 Apr 2014
Date

16 APR 2014
Date

Deb Kashur
Field Inspector's Signature

Charles E. Ehr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:30

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 69
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 06/21/2013
Slope 734
Intercept -865305

IR Calibration Date: 06/21/2013
 3um 9um

0th Coef(*100): -21281 -23057
1st Coef(*100): 292132 137092
2nd Coef(*100): 2334 1231
H2O adj(mg/l*10k): 497 482

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:44
02 Std. Sol.	0.020	16:45
03 Room Air	0.000	16:45
04 Std. Sol.	0.020	16:46
05 Room Air	0.000	16:46
06 Std. Sol.	0.020	16:47
07 Room Air	0.000	16:48


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 08

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:49
02 Std. Sol.	0.305	16:50
03 Room Air	0.000	16:51
04 Std. Sol.	0.305	16:51
05 Room Air	0.000	16:52
06 Std. Sol.	0.305	16:53
07 Room Air	0.000	16:53

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 13080
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 DK
~~0.080 AC~~

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:56

Test	AC	Time
01 Room Air	0.000	16:57
02 *Subject Test	INT*	16:57
03 Room Air	0.000	16:58

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 69
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:58

Test	AC	Time
01 Diagnostic	OK	16:59
02 Room Air	RFI*	16:59
03 Room Air	0.000	17:00

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 69
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:33
02 Std. Gas	0.080	16:33
03 Room Air	0.000	16:34
04 Std. Gas	0.081	16:34
05 Room Air	0.000	16:34
06 Std. Gas	0.081	16:35
07 Room Air	0.000	16:35

Lot No = 16512080A2
Cyl No = 69
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:37
02 Std. Gas	0.081	16:37
03 Room Air	0.000	16:37
04 Std. Gas	0.081	16:38
05 Room Air	0.000	16:38
06 Std. Gas	0.081	16:39
07 Room Air	0.000	16:39

Lot No = 16512080A2
Cyl No = 69
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 Ae

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
04/16/2014 16:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:40
02 Std. Gas	0.081	16:40
03 Room Air	0.000	16:41
04 Std. Gas	0.081	16:41
05 Room Air	0.000	16:42
06 Std. Gas	0.081	16:42
07 Room Air	0.000	16:43

Lot No = 16512080A2
Cyl No = 69
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC