

INTOXILYZER 8000 INSPECTION

Serial No.: 80-005362 Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u>                               | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3377</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>23411080A1</u>	Cyl No. <u>55</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.083</u>	Test 6 <u>0.083</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

L. Tank pressure, Level 3 Func D Sub G P: 700 psi Reg 720 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

17 APR 12  
Date

Cleodis Elden  
Field Inspector's Signature

18 Apr 2012  
Date

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000          SN 80-005362  
Location = TOXL                    8164.13.00 06/09  
04/17/2012                                  17:31

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            55  
Standard Expiration?        10/01/2013  
Oper No?                    666666

Flow Cal. Date:            08/19/2011  
Slope                        667  
Intercept                    -577009

IR Calibration Date:        08/22/2011  
    3um                        9um

-----  
0th Coef(\*100):            -18658                        -21934  
1st Coef(\*100):            258887                        140528  
2nd Coef(\*100):            2709                            1285  
H2O adj(mg/l\*10k):        593                            548

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:            PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.13.00 06/09  
04/17/2012      17:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:37
02 Std. Sol.	0.018	17:38
03 Room Air	0.000	17:39
04 Std. Sol.	0.018	17:39
05 Room Air	0.000	17:40
06 Std. Sol.	0.019	17:41
07 Room Air	0.000	17:41

08 Sim Temp = 34.0°C

Simul Ser No = 000000  
Std Sol No = 0000  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *Low AC*  
                         *0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.13.00 06/09  
04/17/2012      17:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:57
02 Std. Sol.	0.294	17:58
03 Room Air	0.000	17:59
04 Std. Sol.	0.300	18:00
05 Room Air	0.000	18:00
06 Std. Sol.	0.301	18:01
07 Room Air	0.000	18:02

08 Sim Temp = 34.0°C

Simul Ser No = 000000  
Std Sol No = 0000  
County = 08      Oper No. = 666666

  
\_\_\_\_\_  
Operator Signature  
CHARLES EDER

Remarks:      *LINEARITY TEST*  
                         *0.300 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005362  
Location = TOXL                    8164.13.00 06/09  
04/17/2012                            18:03

Test	AC	Time
01 Room Air	0.000	18:04
02 *Subject Test	INT*	18:04
03 Room Air	0.000	18:05

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 123456  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 000000  
Std Sol No = 0000  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT CHECK*  
*0.100 AC + ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.13.00 06/09  
04/17/2012      18:06

Test	AC	Time
01 Diagnostic	OK	18:07
02 Room Air	RFI*	18:07
03 Room Air	0.000	18:08

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 123456  
Dr. Lic. = ND/DIS821456  
Simul Ser No = 000000  
Std Sol No = 0000  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.13.00 06/09  
04/17/2012      18:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:10
02 Std. Gas	0.082	18:10
03 Room Air	0.000	18:10
04 Std. Gas	0.082	18:11
05 Room Air	0.000	18:11
06 Std. Gas	0.083	18:12
07 Room Air	0.000	18:12

Lot No = 23411080A1  
Cyl No = 55  
Exp Date = 10/01/2013  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.13.00 06/09  
04/17/2012      18:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:13
02 Std. Gas	0.082	18:13
03 Room Air	0.000	18:14
04 Std. Gas	0.082	18:14
05 Room Air	0.000	18:15
06 Std. Gas	0.083	18:15
07 Room Air	0.000	18:16

Lot No = 23411080A1  
Cyl No = 55  
Exp Date = 10/01/2013  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *CALIBRATION CHECK*  
                         *0.080AC*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005362  
Location = TOXL      8164.13.00 06/09  
04/17/2012      18:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:17
02 Std. Gas	0.082	18:17
03 Room Air	0.000	18:17
04 Std. Gas	0.082	18:18
05 Room Air	0.000	18:18
06 Std. Gas	0.082	18:19
07 Room Air	0.000	18:19

Lot No = 23411080A1  
Cyl No = 55  
Exp Date = 10/01/2013  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC