



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (03-2014)

Serial Number <i>20-005938</i>	Instrument Location <i>Dickinson PD</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S). *Completed at State Lab prior to transport*
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>4-17-15 ✓</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>28 Apr 2015</i>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Charles E. Eden*      Certified Date: *17 Nov. 16*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005938  
Location = DKPD                    8164.13.00 06/09  
04/17/2015                            15:12

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=-|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=-|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            14214080A6  
Standard Cyl #?            51  
Standard Expiration?       07/05/2016  
Oper No?                    010107

Flow Cal. Date:            07/15/2013  
Slope                       664  
Intercept                  -578364

IR Calibration Date:       04/06/2015  
    3um                            9um

-----  
0th Coef(\*100):        -20371                    -22227  
1st Coef(\*100):        285505                    135759  
2nd Coef(\*100):        3377                      1435  
H2O adj(mg/l\*10k):    305                        271

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
-----  
Operator Signature  
CHRISTOPHER MESSER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005938  
Location = DKPD      8164.13.00 06/09  
04/17/2015      15:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:21
02 Std. Gas	0.079	15:22
03 Room Air	0.000	15:22
04 Std. Gas	0.079	15:23
05 Room Air	0.000	15:23
06 Std. Gas	0.079	15:23
07 Room Air	0.000	15:24

Lot No = 14214080A6  
Cyl No = 51  
Exp Date = 07/05/2016  
County = 45      Oper No. = 010107

  
Operator Signature  
CHRISTOPHER MESSER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005938  
Location = DKPD      8164.13.00 06/09  
04/17/2015      15:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:36
02 Reference	RFI*	15:36
03 Room Air	0.000	15:37

\*Invalid Test  
Inhibited - RFI

Lot No = 14214080A6  
Cyl No = 51  
Exp Date = 07/05/2016  
County = 45      Oper No. = 010107



Operator Signature  
CHRISTOPHER MESSER

Remarks: